Toxicology Proficiency Testing Program Blood Lead Performance Evaluation

Shipment Date: May 06, 2014

Test Event: 142

PFI: 1067 Jill Taylor, Ph.D. Wadsworth Center - Biggs Laboratory PO Box 509 Empire State Plaza Albany, NY 12201-0509

Report reviewed:			
	Signature	_	Date

Toxicology Proficiency Testing Program
Blood Lead
Shipment Date: May 06, 2014
Test Event: 142

PFI: 1067 Jill Taylor, Ph.D. Wadsworth Center - Biggs Laboratory PO Box 509 Empire State Plaza Albany, NY 12201-0509 Clinical Laboratory Reference System Clinical Laboratory Evaluation Program Wadsworth Center New York State Department of Health

		Your Tarç Response		arget Acceptable Range	Relative Distance from Target				
Analyte	Sample		Target		D/Dmax	-1	0	1	Score
Lead (ug/dL)	BE14-06	8	8	4 - 12	0.00		X		100
- ETAAS-Z	BE14-07	4	5	1 - 9	-0.25		x		100
	BE14-08	28	28	24 - 32	0.00	İ	X	İ	100
	BE14-09	22	22	18 - 26	0.00		х	j	100
	BE14-10	16	17	13 - 21	-0.25		x		100
						'	ı	Analyte	Score: 100
				Average D/Dmax:	-0.10			•	Pass

NOTES:

- 1. Laboratory results were evaluated using criteria specified by CLIA'88 regulations. The analyte blood lead was evaluated against targets established by a greater than or equal to 90% consensus of 20 measurements performed by 18 reference laboratories using reference systems based on atomic absorption spectrometry, inductively coupled plasma mass spectrometry, and bench-top anodic stripping voltammetry methods.
- D/Dmax: D represents the deviation of a result from the target value. Dmax represents the maximal allowable deviation from that target. For satisfactory performance, the D/Dmax value must fall between -1 and +1. A negative D/Dmax indicates that your result is below the target value; a positive D/Dmax means your result is above the target value. A blank entry in this column indicates that your result either contains a qualifier (< or >) or is non-gradable, in which case "N/G" is shown in the "Score" column. The average D/Dmax is provided to assess overall test performance for each analyte. A close review of your laboratory's results is recommended if D/Dmax is > +/- 0.5 for a result or analyte.
- 3. The graph plots show the relative distance of your laboratory's result (represented by an "X") from the target value for each sample analyzed. Any result exceeding the high or low limit by >25% of the D/Dmax is indicated by a pound sign (#).
- 4. Summaries of participant statistics from this and prior Toxicology Blood Lead proficiency test events are available on the Internet at: http://www.wadsworth.org/testing/lead/ptresults.htm

Toxicology Proficiency Testing Program
Blood Lead
Shipment Date: May 06, 2014
Test Event: 142

PFI: 1067 Jill Taylor, Ph.D. Wadsworth Center - Biggs Laboratory PO Box 509 Empire State Plaza Albany, NY 12201-0509 Clinical Laboratory Reference System Clinical Laboratory Evaluation Program Wadsworth Center New York State Department of Health

Event:	May 06, 2014	January 15, 2014	September 11, 2013	
Analyte	Score	Score	Score	Status
Lead (ug/dL)	100% Sat	100% Sat	100% Sat	Successful
Event Score:	100% Sat	100% Sat	100% Sat	Successful

Unsatisfactory (Unsat) performance is the failure to attain the minimum satisfactory score for the category or analyte for a testing event. A second unsatisfactory score in one of the next two testing events for the same analyte or category will result in an unsuccessful performance. Please refer to the CLRS Program Guide, available at www.wadsworth.org/clep, for category specific grading criteria.