

**LABORATORY SCALE TEST**

Before proceeding with this test, review Chapter 2 of the Equipment Handbook. Clean, level, and balance the scale, before starting. Any test weights needed to determine accuracy must have a current Report of Test to Class F tolerances.

**FORM APPROVED OMB NO. 0581-0309.** According to the Paperwork Reduction Act of 1995, an agency may not conduct or sponsor, and a person is not required to respond to a collection of information unless it displays a valid OMB control number. The valid OMB control number for this information collection is 0581-0309. The time required to complete this information collection is estimated to average 10 minutes per response and 3 seconds of recordkeeping, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information.

Date	Field Office	Agency	Location
Scale make, model, and serial No.:		Capacity	x division size (d):

Check mark the Class or Type of use, not both. A marked scale has a roman numeral II or III on the front or the i.d. plate. A marked scale is tested using either Class II or Class III tolerances, regardless of what it is used for.

Class II

Class II	
Number of Divisions	Tol.
0-5,000	1d
5,001-20,000	2d
20,001+	3d

Class III

Class III	
Number of Divisions	Tol.
0-500	1d
501-2,000	2d
2,001-4,000	3d
4,001+	5d

An unmarked scale is tested according to each type of its uses: Precision (and separations from work portions ≤ 100 g), Moisture (and separations from work portions > 100 g), and General (and separations from work portions > 500 g).

Type

- 

Type	d	Test Loads Required	Tol.
Precision, etc.	0.01 g	1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 20, 50, 100 g	0.02 g
Moisture, etc.	0.1	above plus 150, 200, 250, 300, 400 g	0.2
General, etc.	1	above plus 500, 600, 700, 1000, 1500 g	1

1. **Class F weights** used in this exam were last tested on this date (maximum 3 years) \_\_\_\_\_


2. **Zero balance** the scale before the sensitivity test, shift test and increasing load test.

3. **Sensitivity** Mechanical scales only. At zero, add weight equal to one scale division. The scale indicator should move to the top of the trig loop (or readings should change by at least one division). Repeat the test at the scale's maximum test load

4. **Shift Test** Place weight of approx. 1/2 of the maximum test load (applied during the increasing load test) halfway between the center and the edge of each quadrant. Apply the basic tolerance for the type of scale and test load.



	Quadrant 1	Quadrant 2	Quadrant 3	Quadrant 4
Reading				



5. **Increasing Load** For marked scales, test at each load from 1 g to 1500 g (or max. capacity) as indicated below. For unmarked scales test at each load indicated above, for each of the scale's uses. Evaluate the scale according to the tolerances shown above for the type of scale tested.

Test Load	Indication						
1g		7g		100g		500g	
2		8		150		600	
3		9		200		700	
4		10		250		1000	
5		20		300		1500	
6		50		400			

**6. Decreasing Load.**

Remove the test weight removing weights from the increasing load test.


Test Load	Indication
400	
200	
50	

Test Load	Indication
0	

7. **Balance Change.** Record the maximum no-load indication during the above tests. Tolerance is 1d.

Approved  Disapproved

Tested by: \_\_\_\_\_

U.S. DEPARTMENT OF AGRICULTURE FEDERAL GRAIN INSPECTION SERVICE <b>LABORATORY SCALE TEST</b> <i>Before proceeding with this test, review Chapter 2 of the Equipment Handbook. Clean, level, and balance the scale, before starting. Any test weights needed to determine accuracy must have a current Report of Test to Class F tolerances.</i>		<b>FORM APPROVED OMB NO. 0580-0013.</b> According to the Paperwork Reduction Act of 1995, an agency may not conduct or sponsor, and a person is not required to respond to a collection of information unless it displays a valid OMB control number. The valid OMB control number for this information collection is 0580-0013. The time required to complete this information collection is estimated to average 10 minutes per response and 3 seconds of recordkeeping, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing the collection of information.																																																																															
Date <b>1</b>	Field Office <b>2</b>	Agency <b>3</b>	Location <b>4</b>																																																																														
Scale make, model, and serial No.: <b>5</b>		Capacity x division size (d): <b>6</b>																																																																															
<p style="font-size: small;">Check mark the Class or Type of use, not both. A marked scale has a roman numeral II or III on the front or the i.d. plate. A marked scale is tested using either Class II or Class III tolerances, regardless of what it is used for.</p> <div style="display: flex; justify-content: space-around;"> <div style="width: 45%;"> <p>Class II <input type="checkbox"/> <b>7</b></p> <table border="1" style="width:100%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th colspan="2" style="text-align: center;">Class II</th> </tr> <tr> <th style="width: 60%;">Number of Divisions</th> <th style="width: 40%;">Tol.</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">0-5,000</td> <td style="text-align: center;">1d</td> </tr> <tr> <td style="text-align: center;">5,001-20,000</td> <td style="text-align: center;">2d</td> </tr> <tr> <td style="text-align: center;">20,001+</td> <td style="text-align: center;">3d</td> </tr> </tbody> </table> </div> <div style="width: 45%;"> <p>Class III <input type="checkbox"/> <b>7</b></p> <table border="1" style="width:100%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th colspan="2" style="text-align: center;">Class III</th> </tr> <tr> <th style="width: 60%;">Number of Divisions</th> <th style="width: 40%;">Tol.</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">0-500</td> <td style="text-align: center;">1d</td> </tr> <tr> <td style="text-align: center;">501-2,000</td> <td style="text-align: center;">2d</td> </tr> <tr> <td style="text-align: center;">2,001-4,000</td> <td style="text-align: center;">3d</td> </tr> <tr> <td style="text-align: center;">4,001+</td> <td style="text-align: center;">5d</td> </tr> </tbody> </table> </div> </div> <p style="font-size: x-small;">An unmarked scale is tested according to each type of its uses: Precision (and separations from work portions ≤ 100 g), Moisture (and separations from work portions &gt; 100 g), and General (and separations from work portions &gt; 500 g).</p> <div style="display: flex; justify-content: space-between;"> <div style="width: 15%;"> <p>Type <b>8</b></p> <input type="checkbox"/>   <input type="checkbox"/>   <input type="checkbox"/> </div> <table border="1" style="width:85%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th style="width: 15%;">Type</th> <th style="width: 15%;">d</th> <th style="width: 45%;">Test Loads Required</th> <th style="width: 25%;">Tol.</th> </tr> </thead> <tbody> <tr> <td>Precision, etc.</td> <td style="text-align: center;">0.01 g</td> <td style="text-align: center;">1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 20, 50, 100 g</td> <td style="text-align: center;">0.02 g</td> </tr> <tr> <td>Moisture, etc.</td> <td style="text-align: center;">0.1</td> <td style="text-align: center;">above plus 150, 200, 250, 300, 400 g</td> <td style="text-align: center;">0.2</td> </tr> <tr> <td>General, etc.</td> <td style="text-align: center;">1</td> <td style="text-align: center;">above plus 500, 600, 700, 1000, 1500 g</td> <td style="text-align: center;">1</td> </tr> </tbody> </table> </div>				Class II		Number of Divisions	Tol.	0-5,000	1d	5,001-20,000	2d	20,001+	3d	Class III		Number of Divisions	Tol.	0-500	1d	501-2,000	2d	2,001-4,000	3d	4,001+	5d	Type	d	Test Loads Required	Tol.	Precision, etc.	0.01 g	1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 20, 50, 100 g	0.02 g	Moisture, etc.	0.1	above plus 150, 200, 250, 300, 400 g	0.2	General, etc.	1	above plus 500, 600, 700, 1000, 1500 g	1																																								
Class II																																																																																	
Number of Divisions	Tol.																																																																																
0-5,000	1d																																																																																
5,001-20,000	2d																																																																																
20,001+	3d																																																																																
Class III																																																																																	
Number of Divisions	Tol.																																																																																
0-500	1d																																																																																
501-2,000	2d																																																																																
2,001-4,000	3d																																																																																
4,001+	5d																																																																																
Type	d	Test Loads Required	Tol.																																																																														
Precision, etc.	0.01 g	1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 20, 50, 100 g	0.02 g																																																																														
Moisture, etc.	0.1	above plus 150, 200, 250, 300, 400 g	0.2																																																																														
General, etc.	1	above plus 500, 600, 700, 1000, 1500 g	1																																																																														
<p>1. <b>Class F weights</b> used in this exam were last tested on this date (maximum 3 years) _____</p> <p>2. <b>Zero balance</b> the scale before the sensitivity test, shift test and increasing load test.</p> <p>3. <b>Sensitivity</b> Mechanical scales only. At zero, add weight equal to one scale division. The scale indicator should move to the top of the trig loop (or readings should change by at least one division). Repeat the test at the scale's maximum test load</p> <p>4. <b>Shift Test</b> Place weight of approx. 1/2 of the maximum test load (applied during the increasing load test) halfway between the center and the edge of each quadrant. Apply the basic tolerance for the type of scale and test load.</p> <div style="display: flex; justify-content: space-around; align-items: center;"> <div style="width: 40%;"> <p><b>9</b></p> <table border="1" style="width:100%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th style="width: 15%;"></th> <th style="width: 15%;">Quadrant 1</th> <th style="width: 15%;">Quadrant 2</th> <th style="width: 15%;">Quadrant 3</th> <th style="width: 15%;">Quadrant 4</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">Reading</td> <td></td> <td></td> <td></td> <td></td> </tr> </tbody> </table> </div> <div style="width: 10%; text-align: center;">  </div> <div style="width: 10%; text-align: center;">  </div> </div> <p>5. <b>Increasing Load</b> For marked scales, test at each load from 1 g to 1500 g (or max. capacity) as indicated below. For unmarked scales test at each load indicated above, for each of the scale's uses. Evaluate the scale according to the tolerances shown above for the type of scale tested.</p> <table border="1" style="width:100%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th>Test Load</th> <th>Indication</th> <th>Test Load</th> <th>Indication</th> <th>Test Load</th> <th>Indication</th> <th>Test Load</th> <th>Indication</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">1g</td> <td></td> <td style="text-align: center;">7g</td> <td></td> <td style="text-align: center;">100g</td> <td></td> <td style="text-align: center;">500g</td> <td></td> </tr> <tr> <td style="text-align: center;">2</td> <td></td> <td style="text-align: center;">8</td> <td></td> <td style="text-align: center;">150</td> <td></td> <td style="text-align: center;">600</td> <td></td> </tr> <tr> <td style="text-align: center;">3</td> <td style="text-align: center;"><b>10</b></td> <td style="text-align: center;">9</td> <td></td> <td style="text-align: center;">200</td> <td></td> <td style="text-align: center;">700</td> <td></td> </tr> <tr> <td style="text-align: center;">4</td> <td></td> <td style="text-align: center;">10</td> <td></td> <td style="text-align: center;">250</td> <td></td> <td style="text-align: center;">1000</td> <td></td> </tr> <tr> <td style="text-align: center;">5</td> <td></td> <td style="text-align: center;">20</td> <td></td> <td style="text-align: center;">300</td> <td></td> <td style="text-align: center;">1500</td> <td></td> </tr> <tr> <td style="text-align: center;">6</td> <td></td> <td style="text-align: center;">50</td> <td></td> <td style="text-align: center;">400</td> <td></td> <td></td> <td></td> </tr> </tbody> </table> <div style="display: flex; justify-content: space-around; margin-top: 10px;"> <table border="1" style="width: 40%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th>Test Load</th> <th>Indication</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">400</td> <td></td> </tr> <tr> <td style="text-align: center;">200</td> <td style="text-align: center;"><b>11</b></td> </tr> <tr> <td style="text-align: center;">50</td> <td></td> </tr> </tbody> </table> <table border="1" style="width: 40%; border-collapse: collapse; font-size: x-small;"> <thead> <tr> <th>Test Load</th> <th>Indication</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">0</td> <td></td> </tr> </tbody> </table> </div> <p>6. <b>Decreasing Load</b> Perform this test while removing <input type="checkbox"/> weights from the increasing load test.</p> <p>7. <b>Balance Change</b> Record the maximum no-load indication during the above tests. Tolerance is 1d. <b>12</b></p>					Quadrant 1	Quadrant 2	Quadrant 3	Quadrant 4	Reading					Test Load	Indication	Test Load	Indication	Test Load	Indication	Test Load	Indication	1g		7g		100g		500g		2		8		150		600		3	<b>10</b>	9		200		700		4		10		250		1000		5		20		300		1500		6		50		400				Test Load	Indication	400		200	<b>11</b>	50		Test Load	Indication	0	
	Quadrant 1	Quadrant 2	Quadrant 3	Quadrant 4																																																																													
Reading																																																																																	
Test Load	Indication	Test Load	Indication	Test Load	Indication	Test Load	Indication																																																																										
1g		7g		100g		500g																																																																											
2		8		150		600																																																																											
3	<b>10</b>	9		200		700																																																																											
4		10		250		1000																																																																											
5		20		300		1500																																																																											
6		50		400																																																																													
Test Load	Indication																																																																																
400																																																																																	
200	<b>11</b>																																																																																
50																																																																																	
Test Load	Indication																																																																																
0																																																																																	
Approved <input type="checkbox"/> Disapproved <input type="checkbox"/> <b>13</b>		Tested by: <b>14</b>																																																																															

## **Instructins for Completeing FGIS-904**

1. Date the test was performed.
2. Field office that performed or supervised the performance of the test.
3. Agency that performed the test, when applicable.
4. Location of the field office or agency that performed the test.
5. Name of the manufacturer, model number, and scale serial number.
6. Maximum rated scale capacity and the minimum scale division size.
7. Indicate with a check mark if the scale is marked with an accuracy class, such as Class II.
8. Indicate with one or more check marks the application of the unmarked scale; precision, moisture, or general.
9. Record scale indications from the shift test.
10. Increasing Load Test - Test marked scales at each test load indicated up to 1,500 g or capacity, whichever is less. Test unmarked scales at each test load shown in the table checked off in item 8, above.
11. Decreasing Load Test - Test marked scales at each load indicated, as appropriate. Test unmarked scales at approximately  $\frac{1}{2}$  the maximum test load used in the increasing load test. If the scale is mechanical, do not perform the test or complete this portion.
12. Balance Change - Scale indication with zero load at completion of testing must be  $\leq 1$  division.
13. Has the scale met all requirements of the pre-test inspection as well as each (all) of the accuracy tests?
14. Name of the person who performed this test.

Return Form To:

FGIS, Field Management Division

FAX Number (202) 720-1015

Retain in official file.

Any Questions Call: Local FGIS field Office.