

Humanitarian Device Exemption (HDE) Program

Guidance for Industry and Food and Drug Administration Staff

Document issued on September 6, 2019.

The draft of this document was issued on June 13, 2018.

For questions about this document regarding CDRH-regulated devices, contact ORP: Office of Regulatory Programs/DRP1: Division of Submission Support, Center for Devices and Radiological Health, at 301-796-5640.

For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-8010.

This guidance supersedes “Guidance for HDE holders, Institutional Review Boards (IRBs), Clinical Investigators, and Food and Drug Administration Staff, Humanitarian Device Exemptions (HDE) Regulation: Questions and Answers,” issued July 8, 2010.

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. The OMB control number for this information collection is 0910-0661 (expires 09-30-2022).

See additional PRA statement in Section IX of this guidance.

**U.S. Department of Health and Human Services
Food and Drug Administration**



**Center for Devices and Radiological Health
Center for Biologics Evaluation and Research
Office of Orphan Product Development**

Preface

Public Comment

You may submit electronic comments and suggestions at any time for Agency consideration to <https://www.regulations.gov> . Submit written comments to the Dockets Management Staff, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD 20852. Identify all comments with the docket number FDA-2014-D-0223. Comments may not be acted upon by the Agency until the document is next revised or updated.

Additional Copies

CDRH

Additional copies are available from the Internet. You may also send an email request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please use the document number 17040 and complete title of the guidance in the request.

CBER

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), by written request, Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, ocod@fda.hhs.gov or from the Internet at <https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>.

OOPD

Additional copies of this guidance document are also available from the Office of Orphan Products Development (OOPD), Office of Clinical Policy and Programs (OCPP), Food and Drug Administration, 10903 New Hampshire Ave, Silver Spring, MD 20993, or by calling 301-796-8660, or from the Internet at <https://www.fda.gov/industry/developing-products-rare-diseases-conditions/designating-humanitarian-use-device-hud>. You may also send an e-mail request to orphan@fda.hhs.gov.

Table of Contents

I.	Introduction.....	1
II.	Background.....	2
III.	Scope.....	4
IV.	HUD Designations and HDE Applications.....	5
V.	FDA’s Review of HDE Applications	5
A.	HDE Application Required Elements and Filing Review Principles	8
B.	FDA Review Actions for an HDE Application	11
VI.	Assessing Probable Benefit and Risk in an HDE Application	14
VII.	Post-Approval Requirements	18
A.	IRB or Appropriate Local Committee Oversight and Approval.....	18
B.	Adverse Event Reporting.....	19
C.	HDE Supplements.....	20
D.	HDE Periodic Reports.....	20
E.	HUD Designation Re-Evaluation and/or HDE Withdrawal	22
VIII.	Special Considerations for Devices Marketed Under an HDE.....	22
A.	Eligibility for Profit.....	22
B.	The Annual Distribution Number (ADN).....	24
C.	Information to Patients.....	26
D.	HDEs and Pediatric Patients	26
E.	Review and Approval of the Use of HUDs in Clinical Care	27
F.	Review and Approval for Clinical Testing of HUDs.....	30
G.	Emergency Use of HUDs.....	31
IX.	Paperwork Reduction Act of 1995.....	32
	Appendix A – Checklist for Filing Review for HDEs.....	33
	Appendix B – Considerations for the Probable Benefit-Risk Assessment.....	44
	Assessment of Probable Benefit.....	45
	Assessment of Risk.....	47
	Assessment of Probable Benefit-Risk	48
	Appendix C – Probable Benefit-Risk Assessment Summary	53

Humanitarian Device Exemption (HDE) Program

Guidance for Industry and Food and Drug Administration Staff

This guidance represents the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.

I. Introduction

FDA developed this guidance document to provide clarity to industry and FDA staff about the current review practices for the Humanitarian Device Exemption (HDE) Program. This programmatic guidance addresses commonly asked questions about HDEs and Humanitarian Use Devices (HUDs), including FDA actions on HDE applications, post-approval requirements, and special considerations for devices marketed under the HDE Program. This guidance document reflects changes in the HDE Program resulting from statutory amendments made by the 21st Century Cures Act (Cures Act)¹ and explains the criteria FDA considers to determine if “probable benefit” has been demonstrated as part of the Agency’s decision-making process regarding marketing authorization for a HUD. This guidance document also reflects amendments made to the HDE provision of the Federal Food, Drug, and Cosmetic Act (FD&C Act) by the FDA Reauthorization Act of 2017 (FDARA).²

For the purposes of this guidance, “you” refers to the HDE applicant or holder, and “we” refers to FDA. FDA’s guidance documents, including this guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidance means that something is suggested or recommended, but not required.

¹ Pub. L. 114-255.

² Pub. L. 115-52.

II. Background

HUDs are medical devices intended to benefit patients in the treatment or diagnosis of diseases or conditions that affect or are manifested in not more than 8,000 individuals in the United States per year.³ In seeking marketing authorization under an HDE application, the first step is the preparation and submission of a HUD designation request to FDA's Office of Orphan Products Development (OOPD). The HDE application, which is the primary focus of this guidance document, is the second step in seeking marketing authorization for a HUD.

To the extent consistent with the protection of the public health and safety and with ethical standards, the purpose of the HDE provision is to "encourage the discovery and use of devices intended to benefit patients in the treatment and diagnosis of diseases or conditions that affect not more than 8,000 individuals in the United States,"⁴ FDA may grant an HDE, which is an exemption from the effectiveness requirements of sections 514 and 515 of the FD&C Act, if we find that the device meets all of the following criteria:

1. The device will not expose patients to an unreasonable or significant risk of illness or injury, and the probable benefit to health from use of the device outweighs the risk of injury or illness from its use while taking into account the probable risks and benefits of currently available devices or alternative forms of treatment;
2. The device would not be available to a person with the disease or condition in question without the HDE, and no comparable device, other than another device approved under an HDE or Investigational Device Exemption (IDE),⁵ is available to treat or diagnose such disease or condition; and
3. The device is designed to treat or diagnose a disease or condition that affects not more than 8,000 individuals in the United States on an annual basis.⁶

The HDE provision was added to the FD&C Act by the Safe Medical Devices Act of 1990 and included, among other things, a prohibition on profits from sale of HUDs and a requirement that before "use" of a HUD to treat or diagnose patients at a facility, an IRB must approve such use.⁷ For purposes of this guidance, approving the "use" of a HUD (as opposed to approving the "investigational use" or a "clinical investigation" of a device) refers to use of the HUD in the course of routine clinical care to treat or diagnose patients. Subsequent amendments to the FD&C Act have added important flexibility to the HDE program while retaining the purpose of encouraging the discovery of medical devices for use in limited patient populations.

³ 21 CFR 814.3(n). As subsequently explained, the current threshold is "not more than 8,000 individuals in the United States." Before section 3052 of the Cures Act took effect, the threshold was "fewer than 4,000 individuals in the United States."

⁴ Section 520(m)(1) of the FD&C Act.

⁵ An approved IDE permits a device to be shipped lawfully for the purposes of conducting investigations of the device without complying with certain other requirements of the FD&C Act that would apply to devices in commercial distribution. See section 520(g) of the FD&C Act; 21 CFR 812.1(a).

⁶ See section 520(m)(2) of the FD&C Act; 21 CFR 814.104(b)(1)-(3).

⁷ Pub. L. 101-629, section 14.

Contains Nonbinding Recommendations

The Food and Drug Administration Modernization Act of 1997 (FDAMA) included a section on expanding the humanitarian use of devices,⁸ which among other provisions:

- Allowed for the use of HUDs under HDEs without prior IRB approval in certain emergency situations (see Section VIII.G for more information); and
- Provided that FDA may suspend or withdraw an HDE only after providing notice and an opportunity for an informal hearing (see Section VII.E for more information).

The Food and Drug Administration Amendments Act of 2007 (FDAAA) further modified the HDE provision, providing that HUDs indicated for use in pediatric patients, or in a pediatric subpopulation may be sold for profit, subject to certain restrictions.⁹ The scope of HDE-approved devices eligible to make a profit was expanded by section 613 of the Food and Drug Administration Safety and Innovation Act (FDASIA).¹⁰ In any given calendar year, a HUD meeting the statutory criteria can be sold for a profit up until the number of devices sold exceeds the annual distribution number (ADN), which is the number of devices reasonably needed to treat, diagnose, or cure a population of 8,000 individuals in the United States (see Section VIII.A for more information).

FDAAA also added section 515A to the FD&C Act, which requires, among other things, the inclusion of additional information regarding pediatric uses in all original HDE applications, if such information is readily available.¹¹ Specifically, section 515A of the FD&C Act requires that each new HDE application include a description, based on readily available information, of any pediatric subpopulations that suffer from the disease or condition that the device is intended to treat, diagnose, or cure, and the number of affected pediatric patients.¹²

The Cures Act amended the FD&C Act to increase the maximum number of patients affected by the disease or condition that a HUD is designed to treat or diagnose to “not more than 8,000 individuals in the United States.”¹³ Further, the Cures Act removed the requirement that institutional review committees, i.e., IRBs, that supervise the clinical testing of devices or approve the use of HUDs be local.¹⁴ See Sections VIII.E and VIII.F below for more information regarding differences between the “use” of a HUD and clinical investigations involving HUDs. The Cures Act also required FDA to publish a guidance that defines the criteria for establishing “probable benefit” as that term is used in section 520(m)(2)(C) of the FD&C Act.¹⁵ This

⁸ Pub. L. 105-115, section 203.

⁹ Pub. L. 110-85, section 303. Pediatric patients are patients who are younger than 22 years of age at the time of diagnosis or treatment. See section 520(m)(6)(E)(i) of the FD&C Act; 21 CFR 814.3(s).

¹⁰ Pub. L. 112-144. Many of the statutory provisions cited throughout this guidance, including sections 515A(a)(2) and 520(m)(6) of the FD&C Act, were added by section 302 of FDAAA and amended by FDASIA.

¹¹ See Pub L. 110-85, section 302. For further discussion of the information required for pediatric uses of medical devices under section 515A, see the guidance document, “Providing Information About Pediatric Uses of Medical Devices,” (available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/providing-information-about-pediatric-uses-medical-devices>).

¹² See section 515A(a)(2) of the FD&C Act.

¹³ Pub. L. 114-255, section 3052.

¹⁴ Pub. L. 114-255, section 3056.

¹⁵ Pub. L. 114-255, section 3052(b).

Contains Nonbinding Recommendations

guidance includes information to define those criteria for HDE applicants, other stakeholders, and FDA staff. See Section III for more information about the scope of this guidance.

FDARA further amended section 520(m)(4)(B) of the FD&C Act to allow either an IRB or “an appropriate local committee” to approve the use of a HUD to treat or diagnose patients at a facility.¹⁶ We interpret this provision to provide additional flexibility for a healthcare facility to determine the individuals involved in, and processes and procedures used by, the committee that approves the use of HUDs at that facility in order to meet the needs of patients. Note, however, that this FDARA provision did not change the requirements for IRB oversight of a clinical investigation of a HUD. An “appropriate local committee” may not review and approve such a clinical investigation in place of an IRB.¹⁷ See Section VII.A. for additional information regarding IRB and appropriate local committee oversight and approval and Section VIII.F. for additional information regarding investigational use of a HUD.

III. Scope

This guidance provides recommendations to industry and FDA staff about the operational aspects of the HDE Program and also explains the principal criteria that FDA considers when determining if probable benefit(s) to health have been demonstrated for a HUD that is being reviewed through the HDE Program. Additionally, this guidance addresses FDA’s assessment of whether the probable benefit(s) to health from use of the device outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment. The decision tools in Appendices B and C are intended to help staff consider the probable benefit-risk factors discussed in Section VI when reviewing HDE applications. This guidance includes sections on the following topics, among others:

- FDA Review Actions for an HDE Application (Section V.B);
- Assessing Probable Benefit and Risk in an HDE Application¹⁸ (Section VI);
- Post-Approval Requirements (Section VII);
- Special Considerations for Devices Marketed Under an HDE (Section VIII); and
- Appendices to support the HDE Program which include:
 - Filing Checklist (Appendix A)
 - Probable Benefit-Risk Assessment Summaries (Appendices B and C)

The overarching principles in this guidance are applicable to devices that are eligible for review through an HDE application by CDRH as well as devices that are eligible for review through an HDE application by the Center for Biologics Evaluation and Research (CBER). This guidance is not intended to supplant or provide recommendations regarding device-specific data requirements, but it may cover broader areas not addressed in device-specific guidance

¹⁶ Pub. L. 115-52, section 502(b).

¹⁷ See sections 520(g)(3) and 520(m)(4)(A) of the FD&C Act; 21 CFR part 56; 21 CFR part 812.

¹⁸ As required by the Cures Act, this guidance explains the principal criteria that FDA considers when determining if probable benefit(s) to health have been demonstrated for a HUD that is being reviewed through the HDE Program.

documents. In addition, this guidance does not address review issues unique to combination products.

IV. HUD Designations and HDE Applications

Before submitting an HDE application to CDRH or CBER, an HDE applicant must first prepare and submit a HUD designation request to OOPD and receive a HUD designation.¹⁹ For more information on the preparation and submission of a HUD designation request, refer to 21 CFR 814.102(a) and the FDA guidance “[Humanitarian Use Device \(HUD\) Designations](#).”²⁰ In the review of a HUD designation request, FDA will determine whether the device is for a rare disease or condition that affects or is manifested in not more than 8,000 individuals in the United States per year. In the case of a device used for diagnostic purposes, FDA will determine whether the documentation demonstrates that not more than 8,000 individuals per year would be subject to diagnosis by the device in the United States.²¹ After receiving a HUD designation, the HDE applicant may submit an HDE application to the appropriate center. Each applicant must have its own HUD designation to submit an original HDE application for a proposed indication.²² Additionally, the HDE applicant can utilize the HDE pathway only if no other comparable device (other than another device approved under an HDE or under an approved IDE) is available to treat or diagnose the disease or condition.²³

Note that if your device is part of a combination product, an HDE may not be the appropriate pathway to market. For questions about the availability of the HDE pathway for combination products, please contact the Office of Combination Products by email at combination@fda.gov. For questions about Companion Diagnostic Devices, contact CDRH’s Office of Health Technology 7/Office of In Vitro Diagnostics and Radiological Health at oir-policy@fda.hhs.gov.

V. FDA’s Review of HDE Applications

Approval of an HDE application under 21 CFR part 814, Subpart H, is considered “FDA approval” of the device based on, among other criteria, evidence that the device will not expose patients to an unreasonable or significant risk of illness or injury and the probable benefit to health from use of the device outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment.²⁴ In addition, to be eligible for HDE application approval, FDA must determine that

¹⁹ 21 CFR 814.102(a).

²⁰ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/humanitarian-use-device-hud-designations>.

²¹ See 21 CFR 814.102(a)(5).

²² See 21 CFR 814.102(a); 21 CFR 814.104(b)(1).

²³ Section 520(m)(2)(B) of the FD&C Act; 21 CFR 814.104(b)(2).

²⁴ Section 520(m)(2)(C) of the FD&C Act; 21 CFR 814.104(b)(3). As discussed in more detail in Section VI, FDA will perform an assessment of the probable benefits and risks of a device, considering a number of factors, including the target patient population and the size of the population, as well as available alternative treatments or diagnostics, as part of its determination of whether an HDE application meets the statutory standard for approval.

Contains Nonbinding Recommendations

the device would not be available to a person with the disease or condition in question without the HDE application approval and that there is no comparable device, other than another device under an HDE or IDE, available to treat or diagnose the disease or condition.²⁵ A HUD that meets the HDE standard for approval is exempt from the requirement of establishing a reasonable assurance of effectiveness that would otherwise be required under sections 514 and 515 of the FD&C Act.²⁶ FDA approval of an HDE application authorizes an applicant to market a HUD in accordance with approved labeling and indication(s) for use, subject to certain profit and use restrictions set forth in section 520(m) of the FD&C Act.

A HUD under an HDE may not serve as a predicate device for purposes of section 513(i) of the FD&C Act.²⁷ A Premarket Approval application (PMA) could subsequently be submitted for the same device and indication(s) approved under an HDE application if the applicant believes there is sufficient evidence to meet the evidentiary standard of a reasonable assurance of safety and effectiveness; or, if appropriate, the applicant could instead submit a request for classification under section 513(f)(2) of the FD&C Act (a De Novo request). If FDA approves a PMA or grants a De Novo request for the HUD or another comparable device with the same indication, we may withdraw the HDE because the HUD would no longer meet the requirements of section 520(m)(2)(B) of the FD&C Act.²⁸ Section V.A discusses comparable devices, and Section VII.E discusses HUD designation re-evaluation.

FDA's review of HDE applications has similarities to the review of PMA applications, with a few key differences. **Some similarities to the PMA program include:**

- HDE amendments, supplements, and reports are generally subject to similar requirements as those for PMAs (although timeframes differ).²⁹ The requirements for each of these types of HDE submissions refers back to the regulatory requirements for the PMA counterpart.
- HUDs are subject to the quality system (QS) regulation under 21 CFR part 820, and HDE applications must include information in sufficient detail so that FDA can make a knowledgeable judgment about the quality control used in the manufacture of the device.³⁰ Additional information on manufacturing information to include in an HDE application can be found in the FDA guidance, "[Quality System Information for Certain Premarket Application Reviews](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/quality-system-information-certain-premarket-application-reviews)"³¹ If you believe that you cannot comply with or should not be subject to the QS regulation requirements, you may request an exemption or a

²⁵ Section 520(m)(2)(B) of the FD&C Act; 21 CFR 814.104(b)(2).

²⁶ See sections 514, 515, and 520(m) of the FD&C Act.

²⁷ Under 21 CFR 807.92(a)(3), a legally marketed (predicate) device to which a new device may be compared for a determination regarding substantial equivalence is a device that was legally marketed prior to May 28, 1976 (preamendments device), or a device which has been reclassified from class III to class II or I, or a device which has been found to be substantially equivalent through the 510(k) process.

²⁸ 21 CFR 814.118(a).

²⁹ See 21 CFR 814.106 (amendments); 814.108 (supplements); and 814.126 (reports).

³⁰ 21 CFR 814.104(b)(4) and 814.20(b)(4)(v).

³¹ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/quality-system-information-certain-premarket-application-reviews>.

Contains Nonbinding Recommendations

variance from any device QS regulation requirement. Petitions for an exemption or variance must be submitted according to the procedures set forth in 21 CFR 10.30.³²

Key differences between the HDE and PMA programs include the following:

- A HUD under an HDE is *exempt* from the requirement of establishing a reasonable assurance of effectiveness that would otherwise be required under sections 514 and 515 of the FD&C Act.
- HDE applications accepted for filing and to which the applicant does not submit a major amendment are reviewed in 75 days, rather than the traditional 180-day review timeframe for PMA applications.³³
- HDE applications are not subject to user fees.
- For a device approved under an HDE application, medical device reports (MDRs) submitted to FDA in compliance with the requirements of 21 CFR part 803 shall also be submitted to the IRB of record.³⁴ If an appropriate local committee, instead of an IRB, approved the use of the device at a facility, FDA recommends that these MDRs be submitted to that committee. See Section VII below for additional information regarding IRB requirements.
- Use of HUDs in the clinical care of patients at a facility requires approval prior to use by either an IRB or an appropriate local committee, with the exception of emergency use.³⁵ See Section VII below for additional information regarding IRB and appropriate local committee requirements.
- An HDE holder can only make a profit, up until the number of devices sold within the calendar year exceeds the ADN, if the device meets the eligibility criteria for the exemption to the profit prohibition in section 520(m)(6)(A)(i), subject to restrictions in section 520(m)(6) of the FD&C Act. See Section VIII.A below for additional information regarding eligibility for profit.
- The applicant must include in the application a statement that no comparable device, other than another HUD approved under an HDE or a device under an approved IDE, is available to treat or diagnose the disease or condition. The applicant must explain in the HDE application why the device would not be available for the indication in question without the HDE approval.³⁶

Applicants wishing to submit a “modular HDE” for their HDE application may use the procedures outlined in the FDA guidance, “[Premarket Approval Application Modular Review](#).”³⁷ A modular HDE, for purposes of this guidance, is a compilation of sections or “modules” submitted at different times that together become a complete HDE application. HDE applicants

³² See 21 CFR 820.1(e).

³³ 21 CFR 814.40 and 814.114.

³⁴ See 21 CFR 814.126(a).

³⁵ See section 520(m)(4) of the FD&C Act; 21 CFR 814.124(a).

³⁶ See section 520(m)(2)(B) of the FD&C Act and 21 CFR 814.104(b)(2).

³⁷ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/premarket-approval-application-modular-review>.

Contains Nonbinding Recommendations

should include a copy of or reference to FDA's HUD designation letter with each HDE modular submission. Together, the modules are expected to contain the information required for an HDE application.

When submitting an HDE application, applicants must prepare an electronic copy of their submission per the FDA guidance document, "[eCopy Program for Medical Device Submissions](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions),"³⁸ and send the e-copy and cover letter to the current address found on the following websites:

(1) <https://www.fda.gov/cdrhsubmissionaddress>, for devices regulated by CDRH; and

(2) <https://www.fda.gov/about-fda/center-biologics-evaluation-and-research-cber/regulatory-submissions-electronic-and-paper>, for devices regulated by CBER.

A. HDE Application Required Elements and Filing Review Principles

To use this guidance appropriately, HDE applicants and FDA staff should review the following basic principles that are in bold typeface and followed by a description of FDA's review policies and procedures. These principles, and the objective criteria outlined in the Filing Checklist in Appendix A, inform FDA's HDE application filing decisions.

FDA should determine whether the contents of the HDE application allow the substantive review to proceed.

The HDE regulations identify criteria that, if not met, may serve as the basis for FDA refusing to file an HDE application.³⁹ The HDE application must contain the basic administrative and scientific elements listed in 21 CFR 814.104(b), unless the applicant justifies an omission in accordance with 21 CFR 814.104(c). The specific questions in the filing checklist are intended to help FDA ensure that the HDE application contents are sufficiently organized and complete to allow the review team to proceed with a substantive review of the application.

The FDA's filing decision should not be based on a substantive review of the data and information in the HDE application.

The filing review is conducted by FDA to ensure that the HDE application is sufficiently complete and to determine the basic adequacy of the technical elements of the HDE application. Notably, in determining whether an HDE application should be filed, the submitted information should not be evaluated to determine whether the device will expose patients to an unreasonable risk of illness or injury or whether the probable benefit to health from the use of the device

³⁸ FDA has issued guidance to implement section 1136 of FDASIA, which added Section 745A(b) of the FD&C Act ("eCopy Program for Medical Device Submissions," available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/ecopy-program-medical-device-submissions>).

³⁹ 21 CFR 814.112(a).

Contains Nonbinding Recommendations

outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment.

The checklist included in Appendix A is a tool to help ensure that the application contains the necessary information to conduct a substantive review. The elements in the checklist stem from either statutory or regulatory requirements, and the format and content are consistent with the analogous checklists for other types of premarket submissions. FDA generally should not refuse to file an HDE application because we have reviewed the data and believe that the application is ultimately inadequate to meet the standard for HDE approval. Subsequently, the substantive review of the HDE application will evaluate the quality of the content and lead to a decision regarding whether the device will expose patients to an unreasonable or significant risk of illness or injury and whether the probable benefit of the HUD outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment. Concerns identified by the Agency regarding results and outcomes of nonclinical and clinical studies should be addressed in the substantive review and should not preclude a filing decision.

FDA should determine whether the applicant provided a justification for any alternative approach.

If the applicant believes any criteria in the checklist are not applicable, the applicant should explain its rationale. Likewise, the applicant should provide a rationale for any deviation from an applicable device-specific or cross-cutting guidance document and must explain any deviation from FDA-recognized consensus standard.⁴⁰ FDA expects that any item in the checklist that is missing from the application will be addressed with a rationale explaining why the item is not applicable and that any deviations will be explained. If a justification to omit certain information or for taking an alternative approach is provided, FDA will consider the adequacy of that justification or alternative approach during substantive review of the application. A given criterion in the checklist will be considered “Not Present” if the application fails to include either the information requested or a rationale for omission.

HDE application filing reviews.

FDA’s decision to “File” or “Not File” an HDE application should be made in collaboration with the HDE review team and with the appropriate supervisory oversight. FDA will notify the HDE applicant of the filing status within 30 calendar days from the date the HDE application was received.⁴¹ Generally, a small number of missing items from the filing checklist will not preclude a positive filing decision; however, if multiple items are missing such that a substantive review cannot be completed, a “Not File” decision will typically be made. FDA staff may use discretion to determine whether missing items are needed for the application to be sufficiently administratively complete to permit substantive review or to request the missing information interactively during the filing or substantive review.

⁴⁰ 21 CFR 814.104(b)(4) and 21 CFR 814.20(b)(5).

⁴¹ 21 CFR 814.112(a).

Contains Nonbinding Recommendations

Additional considerations when using the filing checklist.

Certain elements of the HDE filing checklist are unique to the HDE Program. These elements are discussed in additional detail below.

HUD Designation

HDE applicants must include a copy of or reference to FDA's HUD designation letter with each HDE application.⁴²

Amount charged for the device

As required by 21 CFR 814.104(b)(5), the applicant must state the amount to be charged for the device, and if the amount is more than \$250 a report or attestation must be provided verifying that the amount charged does not exceed the costs of the device's research, development, fabrication, and distribution. A report must be prepared by an independent certified public accountant, made in accordance with the Statement on Standards for Attestation established by the American Institute of Certified Public Accountants. In lieu of such a report, an applicant may submit an attestation by a responsible individual of the organization, verifying that the amount charged does not exceed the costs of the device's research, development, fabrication, and distribution. If the amount charged is \$250 or less, the requirement for a report by an independent certified public accountant or an attestation by a responsible individual of the organization is waived. If an HDE applicant requests that FDA consider whether the HUD meets the eligibility criteria to qualify for profit making (see Section VIII.A. regarding how to request to make a profit), the applicant must still include this report or attestation in the HDE application.

Comparable Devices

As required by 21 CFR 814.104(b)(2), the applicant must provide a statement that no other comparable device, other than another HUD approved under an HDE or a device under an approved IDE, is available to treat or diagnose the disease or condition. A "comparable device" does not need to be identical to the device that is the subject of the HDE application. However, in applying the "comparable device" exemption criterion, FDA takes into account that the purpose of the HDE Program is to encourage development of devices intended to treat or diagnose diseases or conditions that affect small patient populations. In determining whether a comparable device exists, FDA may consider:

- the device's indications for use and technological characteristics;
- the patient population to be treated or diagnosed with the device; and
- whether the device meets the needs of the identified patient population.

FDA may refuse to file an HDE application if FDA determines that a comparable device is available (other than under another HDE or a device under an approved IDE) to treat or diagnose

⁴² 21 CFR 814.104(b)(1).

such disease or condition for which the approval of the HDE application is being sought.⁴³ FDA cannot approve an HDE application for a HUD if we determine that such a comparable device is available.⁴⁴

B. FDA Review Actions for an HDE Application

After an original HDE application or HDE supplement is accepted for filing and FDA begins its substantive review, the Agency may take the following actions during the course of review:⁴⁵

- Approval order;
- Approvable letter;
- Major deficiency letter;
- Not approvable letter; and
- Denial order.

The review timeframe for original HDE applications and HDE supplements is 75 days.⁴⁶ In addition, if the applicant submits a major amendment, the review timeframe may be extended up to 75 days.⁴⁷ Certain changes to the manufacturing procedure or changes in method of manufacture may qualify to be submitted as a 30-day notice.⁴⁸ For more information regarding 30-day notices, refer to the FDA guidance, “[30-Day Notices, 135-Day Premarket Approval \(PMA\) Supplements and 75-Day Humanitarian Device Exemption \(HDE\) Supplements for Manufacturing Method or Process Changes](#).”⁴⁹ See Section VII.C below for additional details on HDE supplements.

1. Approval Order

FDA will issue an approval order (letter) informing the applicant that the HDE application is approved and that the applicant may begin commercial distribution of the device in accordance with any prescribed conditions of approval after we have completed our review and determined that none of the reasons listed in 21 CFR 814.118 for denying approval applies.⁵⁰

When FDA issues an approval order, the FDA review clock stops. An approval order marks the end of FDA’s review, as it is a final action.

⁴³ See 21 CFR 814.112(a)(2).

⁴⁴ See section 520(m)(2)(B) of the FD&C Act.

⁴⁵ See 21 CFR 814.106, 814.116, and 814.118.

⁴⁶ See 21 CFR 814.108 and 814.114.

⁴⁷ 21 CFR 814.106.

⁴⁸ See 21 CFR 814.108 and 814.39(f).

⁴⁹ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/30-day-notices-135-day-premarket-approval-pma-supplements-and-75-day-humanitarian-device-exemption>.

⁵⁰ See 21 CFR 814.116(b).

2. Approvable Letter

FDA will issue an approvable letter to inform the applicant that we have completed our review of the application and determined that one or both of the following are necessary:⁵¹

- Resolution of minor deficiencies that are identified in the approvable letter. These deficiencies may include, for example, clarifications of previously submitted information, revisions to the labeling, and revisions to or development of a post-approval study protocol.
- Completion of an FDA inspection that finds the manufacturing facilities, methods, and controls in compliance with the QS regulation, 21 CFR part 820, and, if applicable, verifies records pertinent to the HDE application. When this is the case, the approvable letter states that the device is “approvable pending GMP inspection.”

When FDA issues an approvable letter pending resolution of minor deficiencies, we stop the FDA review clock and place the application on hold. When FDA receives a complete response to an approvable letter for an HDE application, we consider it a major amendment and restart the clock with a new 75-day FDA response timeframe.

When FDA issues an approvable letter pending GMP inspection letter, we stop the FDA review clock. Once FDA determines that the GMP issues are resolved, we will issue an approval order if all other minor deficiencies that may have been noted in the approvable letter have also been resolved.

3. Major Deficiency Letter

FDA will issue a major deficiency letter to inform the applicant that the HDE application lacks significant information necessary for FDA to complete our review and request that the applicant amend the application to provide the necessary information regarding the device,⁵² such as:

- Additional clinical experience to demonstrate probable benefit and/or that the device will not expose patients to an unreasonable or significant risk of illness or injury;
- Additional non-clinical data to demonstrate probable benefit and/or that the device will not expose patients to an unreasonable or significant risk of illness or injury (e.g., electromagnetic compatibility, electrical safety, biocompatibility, reliability, software, labeling, animal testing⁵³);
- Scientific rationale for test data provided in the application;
- New validation data and analyses (e.g., due to device modifications made during the course of the HDE application review); or

⁵¹ See 21 CFR 814.116(c).

⁵² See 21 CFR 814.106 and 21 CFR 814.37(b).

⁵³ We support the principles of the “3Rs,” to reduce, refine, and replace animal use in testing when feasible. We encourage applicants to consult with us if they wish to use a non-animal testing method they believe is suitable, adequate, validated, and feasible. We will consider if such an alternative method could be assessed for equivalency to an animal test method.

Contains Nonbinding Recommendations

- A re-analysis of previously submitted data (e.g., alternative analytical method).

When FDA issues a major deficiency letter, we stop the FDA review clock and place the application on hold. When FDA receives a complete response to a major deficiency letter, we consider it a major amendment, restart the clock, and review the amendment within 75 days.

4. Not Approvable Letter

FDA will issue a not approvable letter to inform the applicant that we have completed our review and that we do not believe that the application can be approved ‘as-is’ because of significant deficiencies. The not approvable letter will describe the deficiencies in the application, including each applicable ground for not approving and, where practical, will identify measures required to place the application in an approvable form.⁵⁴

Generally, before FDA issues a not approvable letter, we will first issue a major deficiency letter to provide the applicant with an opportunity to address concerns. However, if an applicant fails to provide an adequate response to a major deficiency letter, or if we have attempted to resolve all deficiencies via interactive review and have not received adequate responses, FDA will issue a not approvable letter.

When FDA issues a not approvable letter, we stop the review clock and place the application on hold. If FDA receives a complete response to a not approvable letter, the amendment will be considered a major amendment, and we restart the clock with a new 75-day FDA response timeframe.⁵⁵

5. Denial Order

FDA may deny approval of an HDE application for any of the reasons identified in 21 CFR 814.118(a). FDA will issue a denial order (letter) when we need to inform the applicant that we have denied approval of the HDE application. The denial order will identify all deficiencies in the application, including each applicable ground for denial and, where practical, will identify measures required to place the application in an approvable form. The denial order will include a notice of an opportunity to request review under section 515(d)(4) of the FD&C Act.⁵⁶

When FDA issues a denial order, we end the FDA review clock if a prior action has not already done so. FDA expects that a denial will normally be preceded by another FDA action that stops the review clock, such as a not approvable letter. However, the FD&C Act does not require any prior FDA action, and FDA may, in appropriate circumstances, proceed directly to issuing a denial order. A denial order marks the end of FDA’s review, as it is considered a final action.

⁵⁴ 21 CFR 814.116(d).

⁵⁵ See 21 CFR 814.44(f)(1) and 814.116(d).

⁵⁶ 21 CFR 814.118(b) and 814.45(b).

6. Acknowledgement of Voluntary Withdrawal

An applicant may, on its own initiative, withdraw an HDE application at any time prior to approval, and for any reason, by submitting an amendment informing FDA of its intent to remove the application from FDA's review.⁵⁷ A voluntary withdrawal action will stop the review clock on the receipt date of the amendment.

Under FDA regulations for review of HDE applications, FDA also considers an original HDE application or HDE supplement to have been voluntarily withdrawn if an applicant fails to respond to a not filing letter, an approvable letter, major deficiency letter, or not approvable letter within 75 days of issuance of the letter.⁵⁸ However, if before the end of that 75-day period, an HDE applicant requests additional time to generate data or provide other information to address the issues identified in the FDA letter, FDA may agree to allow additional time, as appropriate. When additional time is requested, FDA generally would allow up to 360 days to provide a complete response to the FDA action letter. We generally do not find it appropriate to agree to requests for additional time beyond 360 days. FDA intends to notify the applicant with a letter acknowledging voluntary withdrawal of the HDE application or HDE supplement. Any amendment submitted in response to an FDA action letter after FDA's notification acknowledging voluntary withdrawal would be considered a resubmission of the HDE application. As such, it will be assigned a new HDE number and will be subject to the requirements of 21 CFR 814.104.

VI. Assessing Probable Benefit and Risk in an HDE Application

As discussed above, a device that meets the criteria under section 520(m) of the FD&C Act is exempt from the effectiveness requirements of sections 514 and 515 of the FD&C Act.⁵⁹ To approve an HDE application, section 520(m) of the FD&C Act requires that FDA find, among other things, "that the device will not expose patients to an unreasonable or significant risk of illness or injury and the probable benefit to health from the use of the device outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available device[s] or alternative forms of treatment."⁶⁰ To make the required findings, FDA performs an assessment of probable risks and benefits for a device as part of its HDE application review.

⁵⁷ 21 CFR 814.116(e)(3).

⁵⁸ See 21 CFR 814.106 and 814.116(e)(1)-(2). FDA considers a "not filing letter" communicating that an HDE application lacks sufficient information for substantive review to be an FDA request to submit an amendment.

⁵⁹ See sections 514, 515, and 520(m) of the FD&C Act; 21 CFR 814.118(a)(1).

⁶⁰ Section 520(m)(2)(C) of the FD&C Act.

Contains Nonbinding Recommendations

FDA also assesses probable benefits and risks as part of its review of PMAs and De Novo requests,⁶¹ and the Agency has previously presented a benefit-risk framework for benefit-risk determinations in the context of reviewing those applications in the guidance document “[Factors to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications](#).” FDA believes that the benefit-risk framework and factors used to assess PMAs or De Novo requests referenced above is generally appropriate for HDE applications. FDA therefore intends to consider the same factors described in FDA’s benefit-risk framework for evaluating PMAs or De Novo requests when assessing probable benefits and risks for HDE applications.

However, given the different standards and requirements that apply to approval of an HDE application, the weighting of those factors and the nature of the evidence available regarding those factors is likely to differ in the HDE context. Among other differences, the HDE pathway accepts greater uncertainty premarket because a reasonable assurance of effectiveness is not required for a device approved under an HDE application.⁶² Therefore, when compared to a PMA or De Novo request, both of which require a demonstration of reasonable assurance of safety and effectiveness,⁶³ it is anticipated that there will generally be greater uncertainty surrounding the benefit-risk profile based on the evidence submitted in an HDE application.

Moreover, as with the benefit-risk framework for evaluating PMAs or De Novo requests, FDA considers relevant factors as part of the probable benefit-risk assessment for an HDE application in the context of the intended use of the device, including the target patient population and the size of the population. For example, the smaller the patient population for which the device is intended, the greater the uncertainty FDA would typically expect in the review of an HDE application because of the challenges of obtaining clinical data regarding the device. FDA’s probable benefit-risk assessment also takes into account currently available alternative treatments or diagnostics, including their limitations. When available, information characterizing patients’ tolerance for risk, tolerance for uncertainty, and their perspectives on probable benefit may provide useful context during the probable benefit-risk assessment. We encourage applicants to collect and submit patient preference information as available to assist in this assessment. In addition, we encourage patient advocacy groups as well as academicians and other groups to collect patient preference information, which may be collected in conjunction with an applicant for purposes of a future HDE application. Refer to the FDA guidance document, “[Patient Preference Information—Voluntary Submission, Review in Premarket Approval Applications](#),”

⁶¹ See FDA’s guidance, “Factors to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications,” (<https://www.fda.gov/regulatory-information/search-fda-guidance-documents/factors-consider-when-making-benefit-risk-determinations-medical-device-premarket-approval-and-de>) for a discussion of how FDA benefit-risk assessments relate to the Agency’s safety and effectiveness evaluation of a device in the context of reviewing PMAs and De Novo requests.

⁶² Section 520(m)(2) of the FD&C Act.

⁶³ To meet the statutory standard for approval of a PMA, there must be a showing of reasonable assurance that the device is safe and effective. See section 515(d) of the FD&C Act. The De Novo classification process is appropriate for devices that would otherwise be subject to PMA but for which general controls or general and special controls provide a reasonable assurance of safety and effectiveness. See section 513(f)(2) of the FD&C Act.

Contains Nonbinding Recommendations

Humanitarian Device Exemption Applications, and De Novo Requests, and Inclusion in Decision Summaries and Device Labeling.⁶⁴

The Agency's probable benefit-risk framework provides for flexibility and use of scientific judgment in assessing the totality of the evidence to determine if a specific device meets the standard for HDE application approval. This flexibility allows FDA to take into account considerations relevant to HDE applications (e.g., a relatively small patient population) under a framework that is consistent across device marketing submission types. To do so, FDA has developed tools to assist in assessing probable benefit and risk for an HDE application. Refer to the supplementary Considerations for the Probable Benefit-Risk Assessment in Appendix B and the Probable Benefit-Risk Assessment Summary in Appendix C. These tools are intended to reflect differences in probable benefit-risk determinations for an HDE application when compared to other types of device marketing submissions. Note that the tools in this guidance also present questions to consider regarding the factors for the probable benefit-risk assessment.

FDA has also published guidance with respect to making benefit-risk determinations for IDE applications.⁶⁵ However, unlike an IDE, which permits a device to be shipped lawfully for the purpose of conducting a clinical investigation of the device's safety and/or effectiveness, an approved HDE application is a marketing authorization. Accordingly, the two applications have different statutory and regulatory standards and, as a general matter, earlier stages of device development and investigational study under an IDE are typically associated with greater uncertainty than an HDE. Approval of an IDE application to permit investigational use of a device may be appropriate where it is unknown if subjects are likely to benefit from the use of the device, if the risks to the subjects are outweighed by the anticipated benefits to the subjects and the importance of the knowledge to be gained, and the IDE application otherwise satisfies the requirements of 21 CFR part 812.⁶⁶ In contrast, approval of the HDE application, which authorizes the marketing of a device, requires, among other things, a demonstration that there is probable benefit and that the probable benefit outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment.

For HDE applications, probable benefit is present when there is evidence for FDA to reasonably conclude that patients are likely to benefit from the use of the device. The probable benefit-risk decision support tools prompt FDA review staff to consider probable benefit in terms of:

- Type of benefit(s)
- Magnitude of benefit(s);
- Probability of the patient experiencing one or more benefit(s);

⁶⁴ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/patient-preference-information-voluntary-submission-review-premarket-approval-applications>.

⁶⁵ See "Factors to Consider When Making Benefit-Risk Determinations for Medical Device Investigational Device Exemptions," (<https://www.fda.gov/regulatory-information/search-fda-guidance-documents/factors-consider-when-making-benefit-risk-determinations-medical-device-investigational-device>).

⁶⁶ FDA may disapprove or withdraw approval of an IDE application if, among other reasons, "[t]here is reason to believe that the risks to the subjects are not outweighed by the anticipated benefits to the subjects and the importance of the knowledge to be gained" 21 CFR 812.30(b)(4).

Contains Nonbinding Recommendations

- Duration of effect(s);
- Patient perspectives; and/or
- Care-partner (e.g., parent or aide) perspectives.

In addition, the probable benefit-risk decision support tools prompt FDA review staff to consider risk in terms of:

- Severity, types, number and rates of harmful events associated with use of the device;
- Probability of a harmful event;
- Duration of harmful events;
- Risk from false-positive or false-negative results for diagnostics; and
- Patient perspectives; and/or
- Care-partner perspectives.

As with benefit-risk assessments for PMAs and De Novo requests, FDA considers additional factors, including uncertainty and available alternative treatments or diagnostics, as part of assessing whether probable benefits outweigh the probable risks in the context of an HDE application. Sources of evidentiary uncertainty could include, but are not limited to:

- Sample size;
- Duration of follow-up;
- Use of a surrogate outcome; and/or
- Use of non-clinical performance data such as animal testing or computer modeling rather than or in addition to a clinical or surrogate outcome.

The types of evidence that may be used to support approval of an HDE application include investigations using laboratory animals, investigations involving human subjects, nonclinical investigations, and analytical studies for in vitro diagnostics. FDA recognizes that in some instances there may be little or no clinical experience with the device that is the subject of an HDE application. Depending upon the nature of the device and its associated risks, FDA may request that clinical data be collected in support of an HDE application. However, it is also important to recognize that non-clinical data may play a critical role in probable benefit-risk assessments in the context of HDE. Medical devices often have attributes that cannot be tested by clinical methods alone and that play a major role in the performance, safety, or effectiveness of the device. In some cases, non-clinical testing (e.g., engineering performance studies, animal studies, analytical testing, or computer modeling and simulation) can obviate or reduce the need for clinical testing to evaluate certain aspects of device design or performance. Both clinical and non-clinical testing methods may be used to assess the probable benefit (including consideration of its likelihood, magnitude and duration), the probability or severity of a given risk, and/or the success of risk control measures, including risk mitigation measures.

The document at Appendix B, “Considerations for the Probable Benefit-Risk Assessment,” is intended to complement the Probable Benefit-Risk Assessment Summary at Appendix C. FDA staff should use the two tools together while reviewing the HDE application. FDA staff should also refer to this guidance to assist them in making their determinations. HDE applicants may

also consider these tools, but inclusion of these documents as part of an application is not a requirement.

VII. Post-Approval Requirements

HDE applications are subject to a number of post-approval requirements, as described below. In addition to these requirements, a post-approval study (PAS) may be required and described in the approval order.⁶⁷ The guidance document, “[Procedures for Handling Post-Approval Studies Imposed by PMA Order](#),” includes recommendations for PMA post-approval studies.⁶⁸ However, most of the information in this guidance document also applies to post-approval studies imposed for an approved HDE application.

A. IRB or Appropriate Local Committee Oversight and Approval

A HUD with an approved HDE application is approved by FDA for marketing. The HDE holder is responsible for ensuring that a HUD under an approved HDE is administered only in facilities having IRB oversight in accordance with the Agency’s regulation governing IRBs.⁶⁹ In addition, approval by an IRB or an appropriate local committee is required before a HUD under an approved HDE can be used at a facility for clinical care, with the exception of emergency use.⁷⁰ See Section VIII.G., “Emergency Use of HUDs.” Note that we do not interpret the statute to require an IRB or appropriate local committee to review and approve each individual use of a HUD, and the IRB may grant a generalized approval to use the HUD at a facility. In such circumstances, FDA does not require the facility, HDE holder, or practitioner to seek approval from the IRB or appropriate local committee for each use, provided the use of the HUD is within the terms of the generalized approval.

FDA interprets the statutory term “appropriate local committee” to mean a standing committee for the facility that has expertise and experience in reviewing and making treatment decisions for clinical care, particularly in applying innovative medical device technologies to clinical care. As such, a standing committee for the facility that includes physicians with experience in the treatment of rare diseases or conditions would be considered an appropriate local committee by the Agency. Depending on the facility and the charters of its committees, examples of an appropriate local committee may include a peer review committee, a credentialing committee, or a Quality Care Committee. We recommend that the committee include a senior executive level medical staff or faculty member (e.g., the Chief Medical Officer, Physician-in-Chief, Surgeon-in-Chief, Department Chair). In addition, FDA interprets the term “appropriate” to mean that members of the appropriate local committee are free of financial and other conflicts of interest in decisions pertaining to the use of the HUD in clinical care or they recuse themselves from such

⁶⁷ 21 CFR 814.126(a) and 814.82(a)(2).

⁶⁸ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/procedures-handling-post-approval-studies-imposed-pma-order>.

⁶⁹ See section 520(m)(4)(A) of the FD&C Act. FDA regulations governing IRBs are in 21 CFR part 56.

⁷⁰ See section 520(m)(4)(B) of the FD&C Act and 21 CFR 814.124(a).

Contains Nonbinding Recommendations

decisions in which they have a conflict of interest. Merely because a facility has a standing committee does not mean the committee is appropriate to review use of a HUD in clinical care.

The HDE holder is not required to submit the names and addresses of the reviewing IRBs or appropriate local committees to FDA. However, as required in 21 CFR 814.126(b)(2), the HDE holder must maintain records of:

- The names and addresses of the facilities to which the HUD was shipped (which FDA interprets to mean information for facilities to which the HUD was shipped to treat or diagnose patients within the U.S.);
- Correspondence with reviewing IRBs; and
- Any other information requested by a reviewing IRB or FDA.

FDA recommends that HDE holders likewise maintain correspondence with reviewing appropriate local committees as well as other information that such committees may require. Additional information regarding the role of IRBs and appropriate local committees with respect to HUDs is available in Section VIII.E, below.

B. Adverse Event Reporting

All adverse events, whether expected or not, must be reported and evaluated in accordance with Medical Device Reporting requirements in 21 CFR part 803. Device manufacturers and user facilities are required to submit medical device reports to FDA and to the “IRB of record” (i.e., the IRB that oversees use of the HUD at the facility where the adverse event occurred) after HDE approval.⁷¹ In the event that an appropriate local committee approved the use of the HUD for routine clinical care at that facility, instead of an IRB, we recommend that manufacturers submit MDRs to that committee.

Accordingly, manufacturers must submit MDRs to FDA and the IRB of record (with submission to the appropriate local committee that approved the use of the HUD recommended, if applicable) when they become aware of information reasonably suggesting that a HUD may have caused or contributed to a death or serious injury, or has malfunctioned and would be likely to cause or contribute to a death or serious injury if the malfunction were to recur.⁷²

User facilities must submit MDRs to FDA, the IRB of record (with submission to the appropriate local committee that approved the use of the HUD recommended, if applicable), and, if known, the manufacturer when they become aware of information reasonably suggesting that a HUD may have caused or contributed to the death of a patient, and must submit reports to the manufacturer (or to FDA and the IRB of record if the manufacturer is unknown) when they become aware of information reasonably suggesting that a HUD may have caused or contributed

⁷¹ See section 519(a) and (b) of the FD&C Act; 21 CFR 803.30, 803.50, and 814.126(a). See 21 CFR 803.3(d) for the definition of a device user facility.

⁷² See 21 CFR 803.50 and 814.126(a).

Contains Nonbinding Recommendations

to a serious injury to a patient of the facility.⁷³ As defined by 21 CFR 803.3(w), a serious injury means an injury or illness that:

- Is life-threatening;
- Results in permanent impairment of a body function or permanent damage to a body structure; or
- Necessitates medical or surgical intervention to preclude permanent impairment of a body function or permanent damage to a body structure.

Adverse events for HUDs that are approved and labeled for pediatric patients or in a pediatric subpopulation, as described in section 520(m)(6)(A)(i)(I) of the FD&C Act, and exempt from the profit prohibition will be reviewed periodically by FDA's Pediatric Advisory Committee (PAC).⁷⁴ Additional information on upcoming PAC meetings is available on the FDA website.⁷⁵

If the HUD is being investigated in a clinical study under an IDE, adverse events that occur during the study should be reported in accordance with 21 CFR 812.150(a)(1) and 812.150(b)(1).

C. HDE Supplements

After FDA approval of an original HDE application, an applicant shall submit an HDE supplement for review and approval by FDA before making a change affecting the safety or probable benefit of the device.⁷⁶

If you wish to request new indications for use for a device under an HDE (e.g., for a different disease or condition) you must obtain a new HUD designation and submit a new original HDE application in compliance with 21 CFR 814.110. If you are submitting a new original HDE application, please contact OOPD to discuss obtaining a new HUD designation. In the new HDE application, any relevant information or data submitted in the HDE application for the original indication may be incorporated by reference.

D. HDE Periodic Reports

You must submit periodic reports for HDEs in accordance with the approval order under 21 CFR 814.126(b). HDE periodic reports must include the following information unless FDA specifies otherwise:

- An update of the information required under 21 CFR 814.102(a) to demonstrate that the HUD designation is still valid.⁷⁷ An updated annual incidence reassessment (AIR) based on updated numbers to show that the target population for the disease or condition for which the device has been designated is not more than 8,000 per year provides this information. The AIR refers to the HUD designated population, which in some cases may

⁷³ 21 CFR 803.30 and 814.126(a).

⁷⁴ Section 520(m)(7) of the FD&C Act.

⁷⁵ <https://www.fda.gov/advisory-committees/committees-and-meeting-materials/pediatric-advisory-committee>

⁷⁶ 21 CFR 814.108 and 814.39.

⁷⁷ See 21 CFR 814.126(b)(1)(i).

Contains Nonbinding Recommendations

be larger than the approved indication under the HDE (i.e., if the HDE approval covers only a certain indication within the designated disease or condition). In reviewing this information, the reviewing center, CDRH or CBER, may refer the AIR to OOPD for further evaluation if necessary.

- An update to the explanation of why the device would not be available unless an HDE were granted, a statement that no other comparable device (other than another HUD under an HDE or a device under an approved IDE) is available to treat or diagnose the disease or condition, and an updated discussion of the risks and benefits of currently available devices or alternative forms of treatment in the United States.⁷⁸ Information on HDE approvals can be found on the publicly accessible search engine on the FDA website.⁷⁹
- An update to the explanation of why the probable benefits to health from the use of the device outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment.⁸⁰
- An update to the amount to be charged for the device and, if over \$250, a report by an independent certified public accountant or an attestation by a responsible individual of the organization verifying that the amount charged does not exceed the costs of the device's research, development, fabrication, and distribution.⁸¹
- The number of devices that have been shipped or sold since initial marketing approval and, if the number shipped or sold exceeds 8,000, an explanation and estimate of the number of devices used per patient.⁸² FDA interprets this regulation to require HDE holders to report the total number of HUDs shipped or sold per year pursuant to the approved HDE application in the U.S., no matter how the HUDs are used (whether for the approved indication(s), emergency use, or otherwise). However, for devices that have both an HDE application approval and a PMA approval for a different indication, you need report only the number of devices that are shipped or sold pursuant to the HDE unless specifically required otherwise by the PMA Approval Order.
- Information describing the clinical experience with the approved device, including safety information that is known or reasonably should be known to the applicant, and any medical device report made under 21 CFR part 803.⁸³
- A summary of any changes made to the device in accordance with supplements submitted under 21 CFR 814.108.⁸⁴

⁷⁸ See 21 CFR 814.126(b)(1)(ii) and 814.104(b)(2).

⁷⁹ <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfHDE/hde.cfm>

⁸⁰ See 21 CFR 814.126(b)(1)(ii) and 814.104(b)(3).

⁸¹ See 21 CFR 814.126(b)(1)(ii) and 814.104(b)(5).

⁸² See 21 CFR 814.126(b)(1)(iii).

⁸³ See 21 CFR 814.126(b)(1)(iv).

⁸⁴ See 21 CFR 814.126(b)(1)(v).

E. HUD Designation Re-Evaluation and/or HDE Withdrawal

If, based on information contained in the HDE periodic reports, FDA is concerned that the HUD designation may no longer apply to your device, we may contact you for additional information, re-evaluate and possibly revoke your HUD designation, and/or withdraw HDE approval.⁸⁵

If we make the determination that more than 8,000 individuals in the United States are affected by or manifest a certain disease or condition per year, we may consider whether your HUD designation should be revoked in accordance with 21 CFR 814.102(c) and your HDE withdrawn in accordance with 21 CFR 814.118. In making this determination, we intend to consider factors such as the number of patients with the disease or condition and the public health need for the device. The investigational use of a HUD outside of the HUD designation would not count toward the limit of 8,000 individuals per year. Per section 520(m)(5) of the FD&C Act, FDA may suspend or withdraw an HDE only after providing notice and an opportunity for an informal hearing. We intend to discuss the regulatory options with the HDE holder before revoking a HUD designation.

If FDA subsequently approves a PMA or grants a De Novo request for the HUD or another comparable device with the same indication(s), we may withdraw the HDE because the HUD would no longer meet the requirements of section 520(m)(2)(B) of the FD&C Act.⁸⁶

VIII. Special Considerations for Devices Marketed Under an HDE

A. Eligibility for Profit

Except in certain circumstances, HUDs under an HDE cannot be sold for an amount that exceeds the costs of research and development, fabrication, and distribution of the device (i.e., for profit).⁸⁷ If a HUD is studied in a clinical investigation for a new indication, the sponsor of the clinical investigation cannot charge subjects or investigators a price higher than necessary to recover the costs of manufacture, research, development, and handling.⁸⁸ Any costs for which a subject in a clinical investigation is responsible must, when appropriate, be provided in the informed consent document.⁸⁹

Under section 520(m)(6)(A)(i) of the FD&C Act, as amended by FDASIA, a HUD under an HDE is only eligible to be sold for profit if the device meets the following criteria (i.e., the eligibility criteria):

⁸⁵ See 21 CFR 814.102(c), 814.118(a)(9), and 814.126(b)(1).

⁸⁶ See 21 CFR 814.118(a).

⁸⁷ Section 520(m)(3) of the FD&C Act.

⁸⁸ See 21 CFR 812.7(b).

⁸⁹ 21 CFR 50.25(b)(3).

Contains Nonbinding Recommendations

- The device is intended for the treatment or diagnosis of a disease or condition that occurs in pediatric patients or in a pediatric subpopulation, and such device is labeled for use in pediatric patients or in a pediatric subpopulation in which the disease or condition occurs; or
- The device is intended for the treatment or diagnosis of a disease or condition that does not occur in pediatric patients or that occurs in pediatric patients in such numbers that the development of the device for such patients is impossible, highly impracticable, or unsafe.

If a device under an HDE does not meet either of the eligibility criteria, the device cannot be sold for profit. FDA reviews the financial information in the HDE holder's initial application and periodically thereafter. When approving the use of a HUD for treatment or diagnosis of patients in clinical care, the IRB or appropriate local committee is not required to review a justification for the amount charged for the HUD. For the purposes of this guidance, the descriptions below are intended to provide additional clarity regarding each component of the eligibility criteria.

Occurs in pediatric patients or in a pediatric subpopulation – This would be a disease or condition that can occur in patients who are younger than 22 years of age.

Does not occur in pediatric patients – This would be a disease or condition that occurs only in patients who are 22 years of age or older. An example of a disease that does not occur in pediatric patients is Alzheimer's disease.

Impossible or highly impracticable – When determining whether the development of a HUD in pediatric patients is "impossible" or "highly impracticable," FDA considers information provided by the applicant to FDA, including publicly-available information such as published literature, which demonstrates that the sponsor would not be able to conduct the necessary clinical investigation(s) in the pediatric population for the device.

For example, FDA may determine that the development of a particular device is "impossible" or "highly impracticable" in pediatric patients if the device is intended to treat a disease or condition that has a pediatric annual incidence that is so small, or if the prevalence of the pediatric patients living with the disease is so small, or if the pediatric population is so geographically dispersed to prevent sufficient patient recruitment in the pediatric population for a clinical investigation. Because of the speed and efficiency of modern communications tools, geographic dispersion will typically justify a determination that development is impossible or highly impracticable only in extraordinary circumstances and will generally have to be coupled with very small population size. FDA does not consider economic factors (such as the costs associated with conducting a clinical investigation) as a basis for being "impossible" or "highly impracticable."

Unsafe – FDA may determine that the development of a HUD in pediatric patients is "unsafe" if the applicant has provided information, including publicly-available information such as published literature, to FDA that demonstrates that the device would expose pediatric patients to an unreasonable or significant risk of illness or injury. If FDA determines that the HUD is eligible to be sold for profit because development of the

Contains Nonbinding Recommendations

device in pediatric patients would be “unsafe,” the labeling (e.g., warnings or contraindications) for the device should reflect the safety concern.

An HDE applicant whose device meets one of the eligibility criteria and who wishes to sell its HUD for profit should provide adequate supporting documentation to FDA in its original HDE application to demonstrate to FDA that the HUD meets the eligibility criteria. An HDE holder whose HDE application was approved prior to the enactment of FDASIA on July 9, 2012, and who wishes to sell its HUD for profit should provide adequate supporting documentation to FDA in an HDE supplement to demonstrate to FDA that the HUD meets the eligibility criteria. If FDA determines that the HUD meets the eligibility criteria, FDA will then determine the ADN for the HUD when FDA approves the HDE application or supplement.⁹⁰

B. The Annual Distribution Number (ADN)

Under section 520(m)(6) of the FD&C Act, if FDA makes a determination that a HUD meets the eligibility criteria, in any given calendar year, a HUD can be sold for a profit up until the number of devices sold exceeds the annual distribution number (ADN).

The ADN is determined by FDA:

- When the Agency approves the original HDE application; or
- When the Agency approves an HDE supplement for an HDE application approved before the enactment of FDASIA on July 9, 2012, if the HDE holder seeks a “determination” for the HUD in an HDE supplement based upon the profit-making eligibility criteria, and FDA determines that the HUD meets the eligibility criteria.⁹¹

Under section 520(m)(6)(A)(ii) of the FD&C Act, the ADN is defined, with respect to a device under an HDE, as the number of devices “reasonably needed to treat, diagnose, or cure a population of 8,000 individuals in the United States.” The applicant should provide the number of devices per year reasonably needed for each individual in the HDE application or HDE supplement and provide adequate documentation to support such number in order to provide a basis for FDA to calculate the ADN.

When determining the ADN, FDA considers the number of devices per year reasonably needed to treat, diagnose, or cure an individual (“first multiplier”) and multiplies that value by 8,000 (“second multiplier”). By law, the second multiplier is always 8,000. Therefore, the ADN will be equal to or greater than 8,000, depending on the value of the first multiplier. For example, the target population estimate for the intended use may be 3,000 individuals, but if 2 devices are reasonably needed per year to treat, diagnose or cure a patient, the ADN would be 16,000 (i.e., 2 multiplied by 8,000 because the second multiplier for the ADN is always 8,000, regardless of the actual population estimate). After FDA has determined the ADN, an HDE holder may submit an

⁹⁰ See section 613(b) of FDASIA and Section VIII.B. for more discussion on the ADN.

⁹¹ See section 520(m)(6)(A)(ii) of the FD&C Act and section 613(b) of FDASIA.

Contains Nonbinding Recommendations

HDE supplement requesting that FDA modify the ADN based upon additional information, and FDA may modify the number.⁹²

As required under 21 CFR 814.126(b)(1)(iii), the HDE applicant is responsible for monitoring how many devices are shipped or sold in the U.S. each year, and if that number exceeds 8,000, to provide an explanation and estimate to FDA of how the device is being used by patients. Similarly, the HDE holder is responsible for monitoring when the number of devices shipped or sold in a year exceeds the ADN, when the HUDs are approved by FDA to make a profit.⁹³ An IRB or appropriate local committee is not responsible for monitoring the number of uses per year of the HUD.

If the HDE holder ships multiple sizes of a device to help ensure that one of the devices is the appropriate size for the patient(s) when used, it would not be necessary to count all of these devices toward the ADN tally if the additional sizes of the devices (that did not properly fit the patient(s)) are returned to the HDE holder. Unused devices should be returned to the HDE holder to appropriately account for them. The HDE holder should document in its periodic report how many devices are returned to the HDE holder if multiple sizes are shipped. Additionally, HDE holders should keep in mind that if they distribute devices in excess of the ADN, they will not be able to make a profit on those devices.

HDE holders assigned an ADN must immediately notify the Agency if the number of devices distributed in a calendar year exceeds the ADN.⁹⁴ HDE holders should make this immediate notification to the Agency by submitting an HDE report whenever the number of HUDs shipped or sold in a calendar year, however the HUD is used, exceeds the ADN. The statutory notification requirement is generally consistent with the reporting requirement in 21 CFR 814.126(b)(1)(iii) concerning the number of devices shipped or sold regardless of their ultimate use (even if outside their approved indications). However, the statutory provision requires immediate notification when the number shipped or sold in a calendar year exceeds the ADN, whereas the current HDE regulations require periodic reports on a timeframe specified in the HDE approval order.

Once this notification occurs, or once FDA discovers through an inspection that the ADN has been exceeded, then the sales of the HUD for the remainder of the year are subject to the general prohibition on profit (unless FDA approves an ADN modification request in an HDE supplement), and the amount charged for the device must not exceed the cost of research and development, fabrication, and distribution of the device.⁹⁵

In those cases in which a device is approved for a certain indication under an HDE application and is approved under a PMA or has a De Novo request granted for a different indication, sales or shipments of the device pursuant to the PMA or the De Novo request are not subject to the ADN reporting requirement. The ADN relates only to those devices that are marketed under an HDE. Therefore, the manufacturer is required to notify FDA only when sales or shipments

⁹² See section 520(m)(6)(C) of the FD&C Act.

⁹³ See section 520(m)(6)(A)(iii) of the FD&C Act.

⁹⁴ See section 520(m)(6)(A)(iii) of the FD&C Act.

⁹⁵ See section 520(m)(6)(D) of the FD&C Act.

pursuant to the HDE exceed the ADN. If a manufacturer must report the number of sales or shipments of a device approved for certain indications under a PMA, the manufacturer would be responsible for separately reporting sales or shipments of devices marketed for different indications under an HDE per 21 CFR 814.126(b)(1)(iii).

C. Information to Patients

Neither the FD&C Act nor FDA regulations require informed consent from patients who are treated or diagnosed with an HDE-approved HUD in the course of their clinical care. An IRB or appropriate local committee may, however, choose to require that patients receive certain information about the HUD when the committee approves use of the HUD for clinical care at a facility. If a committee requires that patients receive a written document prior to use of the HUD in clinical care, the document should include much of the information found in the HDE patient labeling. If no patient information packet is available, the HDE holder may consider including the following in any written information provided to patients: an explanation that the HUD is designed to diagnose or treat the disease or condition described in the HDE labeling and that no comparable device is available to treat the disease or condition; a description of any ancillary procedures associated with the use of the HUD; a description of the use of the HUD; all known risks or discomforts; and an explanation of the postulated mechanism of action of the HUD in relation to the disease or condition. The IRB or appropriate local committee may decide to require inclusion of this or other information explicitly as part of a written document provided to patients.

The labeling for a HUD approved under an HDE, including any labeling provided to patients, must be truthful and non-misleading.⁹⁶ The device labeling must also include the following statement clarifying that effectiveness has not been demonstrated: “Humanitarian Device. Authorized by Federal law for use in the [treatment or diagnosis] of [specify disease or condition]. The effectiveness of this device for this use has not been demonstrated.”⁹⁷ Additional labeling requirements appear under 21 CFR 814.20(b)(10).

D. HDEs and Pediatric Patients

As discussed above, under section 520(m)(6)(A)(i)(I) of the FD&C Act, a HUD is eligible to be sold for profit if, among other things, the device “is intended for the treatment or diagnosis of a disease or condition that occurs in pediatric patients or in a pediatric subpopulation, and such device is labeled for use in pediatric patients or in a pediatric subpopulation in which the disease or condition occurs.” This provision permits HDE holders to receive a profit from the sale of HUDs that are indicated and labeled for pediatric use, subject to the limit of the ADN.

HUDs marketed under an HDE may be indicated and labeled for pediatric use only or for use in both pediatric and adult patients. Devices that are intended to treat both a pediatric population and an adult population may be included in a single HDE application, but the indications for use should specify use in pediatric patients, or pediatric subpopulation(s), as well as use in adults. In

⁹⁶ See section 502(a) of the FD&C Act, 21 U.S.C. 352(a).

⁹⁷ 21 CFR 814.104(b)(4)(ii).

some cases, the probable benefit-risk profile for devices intended for use in a pediatric population, or in a pediatric subpopulation, may differ from its profile when intended for use in an adult population. Therefore, we recommend that HDE applications for devices intended for use in pediatric populations and in adult populations include data supporting the use in both pediatric and adult populations or an appropriate rationale specifically addressing how the data provided for one population (e.g., adults) are sufficient to support approval of an HDE application with indications for use in both populations. For more information about extrapolating data, refer to the FDA guidance, “[Leveraging Existing Clinical Data for Extrapolation to Pediatric Uses of Medical Devices](#).”⁹⁸

As defined in section 520(m)(6)(E)(i) of the FD&C Act, pediatric patients for purposes of section 520(m) of the FD&C Act are patients who are 21 years of age or younger (i.e., up to, but not including, the 22nd birthday) at the time of the diagnosis or treatment.⁹⁹ As defined by section 520(m)(6)(E)(ii) of the FD&C Act, “pediatric subpopulation” means one of the following populations: neonates, infants, children, or adolescents. Additional information about the definition of pediatric patients and pediatric use as it relates to medical devices can be found in the FDA guidance, “[Premarket Assessment of Pediatric Medical Devices](#).”¹⁰⁰

HUDs that are approved and labeled for pediatric patients or in a pediatric subpopulation as described in section 520(m)(6)(A)(i)(I) of the FD&C Act are required, under section 520(m)(8) of the FD&C Act, to be reviewed annually by FDA’s PAC.¹⁰¹ The PAC annually reviews these HUDs to ensure that the HDE remains appropriate for the pediatric populations for which it was approved. The PAC also conducts periodic review of adverse events for these devices when they are exempt from the profit prohibition.¹⁰²

E. Review and Approval of the Use of HUDs in Clinical Care

As summarized above, an IRB or appropriate local committee must approve the use of a HUD at a given facility before it can be used at that facility.¹⁰³ Therefore, a healthcare professional wishing to use an HDE-approved HUD to treat or diagnose a patient at a facility should obtain approval from the facility’s IRB or the appropriate local committee before use of the HUD, except in certain emergencies where prior approval is not required. See Section VIII.G., “Emergency Use of HUDs.” In reviewing the use of the HUD in clinical care, the IRB or appropriate local committee should be cognizant that FDA has made a determination that the

⁹⁸ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/leveraging-existing-clinical-data-extrapolation-pediatric-uses-medical-devices>.

⁹⁹ See also 21 CFR 814.3(s).

¹⁰⁰ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/premarket-assessment-pediatric-medical-devices>.

¹⁰¹ For more information on the PAC, see <https://www.fda.gov/advisory-committees/committees-and-meeting-materials/pediatric-advisory-committee>.

¹⁰² See section 520(m)(7) of the FD&C Act.

¹⁰³ See section 520(m)(4)(B) of the FD&C Act.

Contains Nonbinding Recommendations

probable benefits to health outweigh the probable risks for use of the HUD only within its approved indication(s). The HDE holder is responsible for ensuring that the HUD is administered in facilities that have oversight by an IRB constituted and functioning in accordance with 21 CFR part 56.¹⁰⁴ Note that an IRB's or appropriate local committee's approval for the "use" of a HUD at a facility to treat or diagnose patients in the course of providing clinical care does not mean that there has been IRB approval of a clinical investigation involving the HUD.

The IRB or appropriate local committee is not required to review and approve each individual use of a HUD, nor is it required to audit medical records of patients who receive a HUD. Rather, the IRB or appropriate local committee may use its discretion to determine how to approve use of a HUD, including consideration of professionals' qualifications through training and expertise to use the device.¹⁰⁵ For example, with the input of members with the appropriate expertise in the clinical area, an IRB or appropriate local committee may specify limitations on the use of the device based upon one or more measures of disease progression, prior use and failure of any alternative treatment modalities, reporting requirements to the committee or committee chairperson, appropriate follow-up precautions and evaluations, or other criteria the committee determines to be appropriate.

1. Process and Considerations for Reviewing the Use of HUDs in Clinical Care

For initial review of a HUD, the IRB or appropriate local committee should perform its review at a convened meeting of the committee.¹⁰⁶ The IRB or appropriate local committee should have policies and procedures in place for the receipt and evaluation of the materials necessary for initial approval and continuing review of the HUD's use at that facility. The policies and procedures should also specify whether the committee requires a consent document for the use of the HUD at that facility.

FDA recommends that the IRB or appropriate local committee follow the review criteria in 21 CFR 56.111 and elsewhere in part 56, where applicable. For example, the IRB or appropriate local committee should review the risks to patients that are found in the HDE-approved product labeling, ensure the risks are minimized, and evaluate whether the risks are reasonable in relation to the proposed use of the device at the facility. FDA also recommends that the IRB or appropriate local committee review the following materials, as applicable, during initial review of a request to use a HUD:

- A copy of the HDE approval order;
- A description of the device;
- The product labeling;

¹⁰⁴ See section 520(m)(4)(A) of the FD&C Act and 21 CFR 814.124(a).

¹⁰⁵ For many HDE-approved HUDs, the HDE holder is required to provide training on the use of the device prior to the healthcare professional using the device. Such requirements would be specified in the HDE application approval order. See 21 CFR 814.126(a) and 814.82(a).

¹⁰⁶ See 21 CFR 56.108, which describes a convened meeting of an IRB for purposes of reviewing FDA-regulated clinical investigations.

Contains Nonbinding Recommendations

- The patient information packet that may accompany the HUD;
- A sample consent form for the use of the HUD in clinical care, if required by the IRB or appropriate local committee or by facility policy; and
- A summary of how the physician proposes to use the device, including a description of any screening procedures, the HUD procedure, and any patient follow-up visits, tests or procedures.

A list of approved HDE applications may be found at <https://www.fda.gov/medical-devices/device-approvals-denials-and-clearances/hde-approvals>. The approval order, labeling, and patient information may be found by selecting the submission number of the appropriate HDE application.

FDA does not require submission of a protocol to the IRB or appropriate local committee for review when the committee is evaluating a request to use the HUD in the clinical care of patients at a facility. However, the IRB, appropriate local committee, or institution may require one under its own policies and procedures.

In addition, FDA does not require committees other than the IRB or appropriate local committee to approve the use of a HUD. However, the institution may require additional review. For example, the use of another committee to provide assessments of specific risks posed by the technology or software compatibility may supplement the IRB or appropriate local committee review.

If a physician wants to use a HUD outside its approved indication(s), FDA recommends that the physician follow the IRB or appropriate local committee's requirements for use of a HUD at that facility, which may include separate approval requirements for use outside the approved indication(s). The IRB or appropriate local committee may also require that the physician obtain informed consent¹⁰⁷ from the patient and ensure that reasonable patient protection measures are followed, such as devising schedules to monitor the patient, taking into consideration the patient's specific needs, and the limited information available about the risks and probable benefits of the device. The extent of oversight in these circumstances is up to the IRB or appropriate local committee. As discussed above, MDRs must be submitted to FDA and to the "IRB of record" if the device may have caused or contributed to death or serious injury and for certain malfunctions. If an appropriate local committee approved the use of the HUD at the facility, FDA recommends that MDRs be submitted to that committee.

2. Continuing Review of the Use of HUDs in Clinical Care

Under FDA's current regulations, an IRB that reviews a request to use a HUD at a facility is responsible for initial as well as continuing review of the HUD.¹⁰⁸ When an appropriate local committee conducts such an initial review instead of an IRB, that appropriate local committee

¹⁰⁷ As noted above, "informed consent" required by a facility in the context of clinical care does not refer to informed consent subject to the requirements in FDA's regulations at 21 CFR part 50.

¹⁰⁸ 21 CFR 814.124(a).

Contains Nonbinding Recommendations

should also conduct continuing review of the HUD. For continuing review, an IRB may use an expedited review procedure in which a chairperson or one or more experienced reviewers carries out the review, similar to the expedited review procedure described at 21 CFR 56.110(b). When an IRB conducts the initial review, a facility may decide to utilize an appropriate local committee to conduct continuing review of the use of the HUD in clinical care.

Appropriate local committees may develop their own policies and procedures for continuing review of a HUD and should determine what type of review procedure is appropriate for each HUD. An expedited procedure, such as that described under 21 CFR 56.110, may be appropriate for continuing review because a HUD marketed under an HDE is a legally marketed device, and its use in clinical care does not constitute “research.” An expedited review does not mean a less-than-substantive review. The individual(s) conducting an expedited review for use of a HUD at a facility should thoughtfully consider the risk and benefit information available and any MDRs.

In addition, FDA does not require that the IRB or appropriate local committee serve as a Data Monitoring Committee. The IRB or appropriate local committee may, however, ask the HDE holder for copies of the safety information submitted to FDA in the periodic reports required by 21 CFR 814.126(b)(1). In this way, information that could have a bearing on human safety would be considered at the time of continuing review.

When an IRB or appropriate local committee is deciding whether to approve the use of a HUD for clinical care of patients at a facility, it does not make a Significant Risk/Non-Significant Risk (SR/NSR) determination. As noted above, use of a legally marketed HUD within its HDE-approved indication at a facility to treat or diagnose patients is not a clinical investigation of a device under 21 CFR part 812.

F. Review and Approval for Clinical Testing of HUDs

A clinical investigation of a HUD that requires submission of an IDE application to FDA or is conducted under the abbreviated requirements for NSR devices at 21 CFR 812.2(b) must be approved and supervised by an IRB.¹⁰⁹ Data may be collected in a clinical investigation **for the HDE-approved indication(s)** without an IDE. An approved IDE permits a device to be shipped lawfully for the purposes of conducting investigations of the device without complying with certain other requirements of the FD&C Act that would apply to devices in commercial distribution.¹¹⁰ As long as the HUD is being studied for the indication(s) in its approved labeling, the HUD is not subject to IDE requirements because the HUD is a legally marketed device and therefore can be lawfully shipped without an IDE. However, regardless of the applicability of the IDE regulation at 21 CFR part 812, other FDA regulatory requirements may still apply, including

¹⁰⁹ See 21 CFR 56.103, 812.2(b)(1)(ii), and 812.42.

¹¹⁰ See 21 CFR 812.1.

Contains Nonbinding Recommendations

requirements for IRB review and approval, financial disclosure, informed consent¹¹¹ and, if applicable, additional safeguards for children.¹¹²

If the IRB receives a request to review an investigation to determine safety or effectiveness of the HUD for a different indication than the HDE-approved indication(s), then the IRB should be aware that this type of clinical investigation is subject to the IDE regulations at 21 CFR part 812. If the device is a SR device, the sponsor of the investigation must submit an IDE application to FDA and obtain FDA approval of that application before starting the clinical investigation.¹¹³ A physician who wants to study a HUD may be the sponsor, investigator, or both for the study. In sum, the investigational use of a HUD under these circumstances must be conducted in accordance with 21 CFR parts 812, 50, 54, and 56.¹¹⁴

Significant Risk/Non-Significant Risk Determinations

An IRB does not have to make a SR/NSR determination when it receives a request to review a clinical investigation of a HUD (e.g., collection of safety and effectiveness data) when that clinical investigation concerns the HDE-approved indication(s) only. As noted above, FDA does not consider such investigations to require an IDE under 21 CFR part 812.

For an investigation of the HUD for indications other than the HDE-approved indication(s), the IRB would need to make a SR/NSR determination if that determination has not already been made by FDA.¹¹⁵ In practice, most sponsors have submitted and obtained FDA approval of an IDE application before submitting such investigations of HUDs to IRBs for review, so IRBs have not needed to make the SR/NSR determination (i.e., FDA had already determined the device was a SR device). However, in the event that a sponsor seeks IRB approval for investigational use of a HUD for an indication other than its approved indication(s) without first obtaining a determination from FDA regarding whether the study is a SR or NSR study, then the IRB should make the SR/NSR determination as required in 21 CFR 812.66.

G. Emergency Use of HUDs

If a physician in an emergency situation determines that IRB or appropriate local committee approval for the use of the HUD at the facility cannot be obtained in time to prevent serious harm or death to a patient, a HUD may be used without prior approval. In this situation, the HDE

¹¹¹ Specific requirements for obtaining informed consent from human subjects apply to FDA-regulated clinical investigations. See 21 CFR part 50, subpart B. Note that, in some cases, facilities may have specific requirements for obtaining informed consent for the use of the HDE-approved HUD in the routine clinical care of patients, but these would not be FDA regulatory requirements.

¹¹² See 21 CFR part 56 for IRB requirements; see 21 CFR part 54 for requirements for financial disclosure by clinical investigators; and see 21 CFR part 50 for requirements for the protection of human subjects, including additional safeguards for children.

¹¹³ 21 CFR 812.20(a).

¹¹⁴ Note that 45 CFR part 46 may be applicable to research involving HUDs under certain circumstances. The applicability of those regulations is outside the scope of this guidance.

¹¹⁵ See 21 CFR 812.66.

Contains Nonbinding Recommendations

holder may ship the HUD, based on the physician's certification of the emergent need and representation that the physician will follow the requirements regarding reporting such use to the chairperson of the IRB or appropriate local committee. The physician must provide notification of the use to the chairperson of the IRB or appropriate local committee, and the notification must include the identification of the patient involved, the date of the use, and the reason for the use.¹¹⁶ FDA regulations require that physicians provide such notification to the chairperson of an IRB in writing within 5 days of the emergency use of the device.¹¹⁷ For facilities at which an appropriate local committee reviews the use of HUDs instead of an IRB, FDA recommends that physicians provide the same required notification of the emergency use to the committee in writing and within 5 days.

FDA further recommends that the physician submit a follow-up report on the patient's condition to the HDE holder. The HDE holder is required under 21 CFR 814.126(b) to submit periodic reports, including the applicant's clinical experience with the device and the number of devices shipped or sold in the US.

IX. Paperwork Reduction Act of 1995

This guidance contains information collection provisions that are subject to review by the Office of Management and Budget (OMB) under the Paperwork Reduction Act of 1995 (44 U.S.C. 3501-3520).

The time required to complete this information collection is estimated to average 100 hours per response, including the time to review instructions, search existing data sources, gather the data needed, and complete and review the information collection. Send comments regarding this burden estimate or suggestions for reducing this burden to:

FDA PRA Staff
Office of Operations
Food and Drug Administration
PRASaff@fda.hhs.gov

An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB control number. The OMB control number for this information collection is 0910-0661 (expires 09/30/2022).

¹¹⁶ See section 520(m)(4) of the FD&C Act.

¹¹⁷ 21 CFR 814.124(a).

Appendix A – Checklist for Filing Review for HDEs

(should be completed within 30 days of DCC receipt)

HDE Number: _____ **Date Received:** _____

HUD Number (from OOPD): _____

Device: _____ **Procode:** _____

Company Name/ Address: _____

Contact Name/Phone Numbers: _____

FDA Staff Member Name: _____

Decision: FDA Staff Recommendation: File ____ Not File ____

Within 15 calendar days of receipt of the HDE application, FDA staff should answer the preliminary questions below, which are used as an initial screening of the HDE application. Depending upon the answers to these preliminary questions, the remainder of the filing review may or may not be necessary. If the responses to the preliminary questions and subsequent consultation with FDA staff identified below indicate that the HDE filing review should not continue, the FDA staff member or the CBER regulatory project manager (RPM) should promptly inform the FDA team (including consulting reviewers and management) and notify the requester using proper administrative procedures.

Preliminary Questions		
Answers in the shaded blocks indicate consultation with an identified Center advisor is needed.	Yes	No
<p>1. Is the product a device [per 201(h) of the FD&C Act] or a combination product (per 21 CFR 3.2(e)) with a device constituent part? If it appears not to be a device (per 201(h) of the FD&C Act) or such a combination product (per 21 CFR 3.2(e)), or you are unsure, consult with the CDRH Product Jurisdiction Officer or CBER Product Jurisdiction Officer to determine the appropriate action and inform management. <i>Provide summary of Product Jurisdiction Officer's determination/recommendation/action in the comments section below.</i></p> <p>If the product does not appear to be a device or a combination product with a device constituent part, mark "No."</p> <p><u>NOTE:</u> If the product is a combination product with a device constituent part, it may not be appropriate for review under an HDE. If the product is a combination product, consult with the CDRH Product Jurisdiction Officer (cdrhproductjurisdiction@fda.hhs.gov) or CBER Product Jurisdiction Officer and inform management.</p>		
Comments:		
<p>2. Is there a copy of, or reference to the determination made by the Office of Orphan Product Development that the device qualifies as a HUD? [814.104(b)(1)]</p> <p>If there is no copy of, or reference to the HUD determination, mark "No."</p>		

Contains Nonbinding Recommendations

<p>3. If a Request for Designation (RFD) was submitted for the device and assigned to your center, identify the RFD # and confirm the following:</p> <ul style="list-style-type: none"> Is the device the same (e.g., design, formulation) as that presented in the RFD submission? Are the indications for use for the device identified in the HDE the same as those identified in the RFD submission? <p>If you believe the product or the indications presented in the HDE have changed from the RFD, or you are unsure, consult with the CDRH Product Jurisdiction Officer (cdhrproductjurisdiction@fda.hhs.gov) or CBER Product Jurisdiction Officer to determine the appropriate action and inform management. <i>Provide summary of Product Jurisdiction Officer's determination/recommendation/action in the comments section below.</i></p> <p>If the answer to either question above is no, mark "No."</p>		
Comments:		
<p>4. Is the device eligible for HDE ?</p> <p><u>NOTE:</u> If the device does not appear to be eligible for review through the HDE program because there is a comparable device available (e.g., a predicate device exists, a De Novo request has been granted for a similar device, or an approved PMA exists for a similar device), you should consult with management and the appropriate CDRH or CBER staff during the filing review to determine the appropriate action.</p> <p>If you believe an application is for a device that is eligible for review through the HDE program and an exemption from the effectiveness provisions, you should (1) complete the 510(k) decision tree to document why the device would be found NSE (<i>attach copy</i>) and (2) obtain concurrence from the appropriate CDRH or CBER staff prior to the filing the original HDE.</p>		
Comments:		
<p>5. Is the applicant the subject of an Application Integrity Policy (AIP)? If "Yes", consult with the CDRH Office of Product Evaluation and Quality/Office of Clinical Evidence/Division of Clinical Evidence & Analysis 1 (CDRH/OPEQ/OCEA/DCEA1) or CBER Office of Compliance and Biologics Quality/Division of Inspections and Surveillance/Bioresearch Monitoring Branch (CBER/OCBQ/DIS/BMB), to determine the appropriate action and provide a summary of the discussion/recommendation/action in the comments section below. Check on web at https://www.fda.gov/inspections-compliance-enforcement-and-criminal-investigations/application-integrity-policy/application-integrity-policy-list</p>		
Comments:		

Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)				
Check "Yes" if item is present, "N/A" if it is not needed and "Not Present" if it is not included but needed.				
		Present		Not Present (No)
	<ul style="list-style-type: none"> A "Not Present" answer may result in a "Refuse to File" decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present ("Yes"). An assessment of the rationale will be considered during the review of the application. 	Yes	N/A	
A.	HDE Content			
1.	Are all required sections in English or accompanied with an English translation? ‡	<input type="checkbox"/>		<input type="checkbox"/>
2.	Is there a table of contents? [814.104(b)(4) and 814.20(b)(2)]	<input type="checkbox"/>		<input type="checkbox"/>

‡ Inclusion of information in an HDE application that is not in English and is not accompanied by an English translation is not an independent basis for a "Refuse to File" decision; however, we recommend providing the sections of your HDE application in English (or accompanied with an English translation) in order to avoid significant delay of review of your submission.

Contains Nonbinding Recommendations

Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)						
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.						
		<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 		Present		Not Present (No)
				Yes	N/A	
	3.	HDE / HUD Information				
	a.	Is there an explanation of why the device would not be available unless an HDE was granted? [814.104(b)(2)]		<input type="checkbox"/>		<input type="checkbox"/>
	b.	Is there a statement that no other comparable device, other than another approved HUD under an HDE or a device under an approved IDE, is available to treat or diagnose the disease or condition? [814.104(b)(2)]		<input type="checkbox"/>		<input type="checkbox"/>
	c.	Is there a discussion of the risks and benefits of currently available devices or alternative forms of treatment? [814.104(b)(2)]		<input type="checkbox"/>		<input type="checkbox"/>
	d.	Is there an explanation of why the probable benefit to health from the use of the device outweighs the risk of injury or illness from its use, taking into account the probable risks and benefits of currently available devices or alternative forms of treatment? [814.104(b)(3)]		<input type="checkbox"/>		<input type="checkbox"/>
	e.	Has the amount to be charged for the device been provided, and if greater than \$250.00, is a report or attestation provided verifying that the amount charged does not exceed the costs of the device’s research, development, fabrication, and distribution? [814.104(b)(5)]		<input type="checkbox"/>		<input type="checkbox"/>
	4.	Is a bibliography provided? [814.104(b)(4) and 814.20(b)(8)(i)]		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	a.	Have copies of key articles been provided and are English translations included, if appropriate? ‡ Check “N/A” if applicant includes a statement that upon searching they found no literature related to their device		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	5.	If a device sample has been requested by FDA, has it been provided or if impractical to submit, has the applicant offered alternatives to allow FDA staff to view or access the device? [814.104(b)(4) and 814.20(b)(9)]		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	6.	Is there a summary of the contents of the HDE? [814.104(b)(4) and 814.20(b)(3)]		<input type="checkbox"/>		<input type="checkbox"/>
	7.	Device Characteristics				
	a.	Is a description of the device included? [814.104(b)(4) and 814.20(b)(4)]		<input type="checkbox"/>		<input type="checkbox"/>
	i.	Pictorial representations? [814.104(b)(4) and 814.20(b)(4)(i)]		<input type="checkbox"/>		<input type="checkbox"/>
	ii.	Materials specifications? [814.104(b)(4) and 814.20(b)(4)(ii)]		<input type="checkbox"/>		<input type="checkbox"/>
			<ul style="list-style-type: none"> If there is a color additive present: <ul style="list-style-type: none"> has the color additive been identified by common name and chemical name, and has the amount of each color additive in the formulation by weight percent of the colored component and total amount (e.g., ppm, µg) in the device been provided? 	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	b.	Is a description of the principles of operation of the device (including components) and properties relevant to clinical function present? [814.104(b)(4) and 814.20(b)(4)(iii)-(iv)]		<input type="checkbox"/>		<input type="checkbox"/>

‡ While submission of key articles that are not in English and are not accompanied by an English translation is not an independent basis for a “Refuse to File” decision, we recommend that copies of submitted articles are provided in English (or accompanied with an English translation) in order to avoid significant delay of review of your submission.

Contains Nonbinding Recommendations

<u>Inventory of Organizational and Administrative Elements</u> (Requirements per 21 CFR 814.112 unless otherwise indicated)					
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.					
<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 			Present		Not Present (No)
			Yes	N/A	
8.	Is the Device Manufacturing Section included? [814.104(b)(4) and 814.20(b)(4)(v)] (See the guidance entitled, “Quality System Information for Certain Premarket Application Reviews,” available at https://www.fda.gov/regulatory-information/search-fda-guidance-documents/quality-system-information-certain-premarket-application-reviews)		<input type="checkbox"/>		<input type="checkbox"/>
	a.	Has a description of the methods, facilities, and controls used in the manufacture, processing, packing, storage, and, where appropriate, installation of the device been provided?	<input type="checkbox"/>		<input type="checkbox"/>
9.	<p>The application includes a summary and full study report* for each nonclinical study provided? [814.104(b)(4) and 814.20(b)(6)(i)]</p> <p>Note: the applicant can reference data located in other applications. Check “Yes” if nonclinical data is not provided in the current application but found in another application. State where the data were provided (e.g., modular application, master file).</p> <p>*Full study report includes objective of the test, description of test methods and procedures, study endpoint(s), pre-defined pass/fail criteria, results summary, and discussion of conclusions. In the event that an applicant is appropriately declaring conformity with a voluntary consensus standard that FDA has recognized pursuant to section 514(c) of the FD&C Act to meet applicable requirements, it may not be necessary to submit full test reports with respect to those requirements. Refer to 13(a). See FDA’s guidance “Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices,” available at https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices.</p>		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	a.	Sterilization	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	b.	Biological/Microbiological	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	c.	Immunological	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	d.	Toxicological/Biocompatibility	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	e.	Engineering (stress, wear, etc.)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	f.	Chemistry/Analytical (typically for IVDs)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	g.	Shelf Life	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	h.	Animal Studies	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	i.	Other Essential Laboratory Testing	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
10.	Is a summary of clinical experience and investigation(s) and results provided? [814.104(b)(4)(i)]		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Contains Nonbinding Recommendations

Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)						
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.						
<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 				Present		Not Present (No)
				Yes	N/A	
		a.	Are the final versions of the clinical protocols included? (If performed under IDE, these should be the final FDA-approved versions of the clinical protocols, incorporating any Notices of Changes.)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.	Is a description of study population demographics provided?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	Is a description of adverse events (e.g., adverse reactions, complaints, discontinuations, failures, replacements) provided?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		d.	Have report forms for patients who died or who did not complete the investigation been provided (i.e., to resolve potential bias)? Check “N/A” only if no patients died or were discontinued.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	11.		Are statistical analyses of the clinical investigations provided, if appropriate? [814.104(b)(4)(i)]	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		a.	Are the results of all analyses identified in the protocol provided?	<input type="checkbox"/>		<input type="checkbox"/>
	12.		Has appropriate draft labeling been submitted? [814.104(b)(4) and 814.20(b)(10)]			
		a.	Physician Labeling	<input type="checkbox"/>		<input type="checkbox"/>
		i.	Are indications for use included?	<input type="checkbox"/>		<input type="checkbox"/>
		ii.	Are contraindications, warnings, and precautions included?	<input type="checkbox"/>		<input type="checkbox"/>
		iii.	Are instructions for use included?	<input type="checkbox"/>		<input type="checkbox"/>
		iv.	Does the labeling include the statement: “Humanitarian Device. Authorized by Federal law for use in the [treatment or diagnosis] of [specify disease or condition]. The effectiveness of this device for this use has not been demonstrated” [814.104(b)(4)(ii)]	<input type="checkbox"/>		<input type="checkbox"/>
		b.	Patient Labeling Check Check “N/A” only if the relevant lead Center has previously indicated that patient labeling is not necessary.	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	Technical/Operators Manual, if applicable	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	13.		Statements/Certifications/Declarations of Conformity			
		a.	Does the application utilize voluntary consensus standard(s) (See section 514(c) of the FD&C Act). <i>This includes both FDA-recognized and non-recognized consensus standards. Select “N/A” if the submission does not utilize voluntary consensus standards.</i>	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		i.	If the application cites FDA-recognized voluntary consensus standard(s), does the application include: a Declaration of Conformity (DOC) as outlined in FDA’s guidance “Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices,” available at	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Contains Nonbinding Recommendations

Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)							
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.							
<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 				Present		Not Present (No)	
				Yes	N/A		
				https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices OR if citing general use of a standard, an explanation of any deviation from the standard? ‡ [814.104(b)(4) and 814.20(b)(5)]			
			ii.	If the application cites non-FDA-recognized voluntary consensus standard(s), does the application explain any deviation from the standard? [814.104(b)(4) and 814.20(b)(5)]	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.	Investigator Financial Disclosure For additional information refer to the guidance document “Financial Disclosure by Clinical Investigators”, available at https://www.fda.gov/regulatory-information/search-fda-guidance-documents/financial-disclosure-clinical-investigators . As required by 21 CFR Part 54, has the applicant submitted for each clinical investigator either: 1. A signed and dated Certification Form (3454) or 2. A signed and dated Disclosure Form (3455) Note: the signature should be from a responsible corporate official or representative of the applicant.		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			i.	For a Certification Form (3454): Is the required list of all investigators and subinvestigators attached to the Form?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			ii.	If box 3 of Form 3454 is checked, does the Form include an attachment with the reason(s) why financial disclosure information could not be obtained?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			iii.	For a Disclosure Form (3455): Does the application provide details of the financial arrangements and interests of the investigator(s) or subinvestigator(s), along with a description of any steps taken to minimize potential bias?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	Environmental Assessment under 21 CFR 25.20(n) [814.104(b)(4) and 814.20(b)(11)]		<input type="checkbox"/>		<input type="checkbox"/>
			i.	If claiming a categorical exclusion, information to justify the exclusion, OR	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			ii.	An environmental assessment (<u>ONLY</u> required for devices that present new environmental concerns)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

‡ If citing general use of a FDA-recognized standard or citing a non-FDA recognized voluntary consensus standard, we recommend that the basis of such use, along with the underlying information or data that supports how the standard was used, be included in the application.

Contains Nonbinding Recommendations

Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)						
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.						
<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 				Present		Not Present (No)
				Yes	N/A	
		d.	Did the application include a completed FORM FDA 3674, <i>Certification with Requirements of ClinicalTrials.gov Data Bank?</i> (42 U.S.C. 282(j)(5)(B) and 42 CFR part 11) Note: Enter the NCT number(s) in the Center Tracking System (CTS)	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			Data from FORM FDA 3674 (mark “Yes” for the applicable one):			
		i.	No clinical trials referenced in application.	<input type="checkbox"/>	<input type="checkbox"/>	
		ii.	Requirements are not applicable to referenced clinical trials.	<input type="checkbox"/>	<input type="checkbox"/>	
		iii.	Requirements are applicable and have been met.	<input type="checkbox"/>	<input type="checkbox"/>	
		f.	Statements of Compliance for Clinical Investigations [814.104(b)(4) and 814.20(b)(6)(ii)(A)-(C)] Select “N/A” if the application does not contain any clinical data from investigations (as defined in 21 CFR 812.3(h)). For multicenter clinical investigations involving both United States (US) and outside the United States (OUS) sites, part (i) should be addressed for the US sites and part (ii) should be addressed for the OUS sites. 21 CFR 812.28 applies to all OUS clinical investigations that enroll the first subject on or after February 21, 2019. Please refer to the guidance document entitled “Acceptance of Clinical Data to Support Medical Device Applications and Submissions - Frequently Asked Questions,” available at https://www.fda.gov/regulatory-information/search-fda-guidance-documents/acceptance-clinical-data-support-medical-device-applications-and-submissions-frequently-asked , for more information.	<input type="checkbox"/>	<input type="checkbox"/>	
		i.	For all clinical investigations conducted in the US, the application includes one of the following for each investigation (<i>check all that apply</i>): <input type="checkbox"/> A statement of compliance with 21 CFR parts 50, 56, and 812. <input type="checkbox"/> A brief statement of the reason for noncompliance with 21 CFR parts 50, 56, and 812. <i>Select “N/A” if the clinical investigations were conducted solely OUS.</i>	<input type="checkbox"/>	<input type="checkbox"/>	
		ii.	For all clinical investigations conducted OUS, the application includes one of the following for each investigation (<i>check all that apply</i>): <input type="checkbox"/> A statement that the clinical investigations were conducted in accordance with good clinical practice (GCP) as described in 21 CFR 812.28(a)(1).	<input type="checkbox"/>	<input type="checkbox"/>	

Contains Nonbinding Recommendations

Inventory of Organizational and Administrative Elements (Requirements per 21 CFR 814.112 unless otherwise indicated)							
Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.							
<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 				Present		Not Present (No)	
				Yes	N/A		
			<input type="checkbox"/> A brief statement of the reason for not conducting the investigation in accordance with GCP and a description of steps taken to ensure that the data and results are credible and accurate and that the rights, safety, and well-being of subjects have been adequately protected. <input type="checkbox"/> A waiver request in accordance with 21 CFR 812.28(c). <i>Select “N/A” if the clinical investigations were conducted solely inside the US.</i>				
	14.	Pediatric Use - Per 515A(a)(2) of the FD&C Act, did the application include, if readily available: [814.104(b)(6) and 814.20(b)(13)]					
		a.	A description of any pediatric subpopulations that suffer from the disease or condition that the device is intended to treat, diagnose, or cure, or statement that no pediatric subpopulation exists for the disease or condition for which the device is intended. This statement does not mean the device is indicated for treating pediatric patients. For additional information refer to the guidance document “Providing Information about Pediatric Uses of Medical Devices - Guidance for Industry and Food and Drug Administration Staff”, available at https://www.fda.gov/regulatory-information/search-fda-guidance-documents/providing-information-about-pediatric-uses-medical-devices .		<input type="checkbox"/>		<input type="checkbox"/>
		b.	The number of affected pediatric patients.		<input type="checkbox"/>		<input type="checkbox"/>
	B.	Issues Identified by FDA Prior to receipt of the HDE Application - history of the applicant with this device [814.104(b)(4) and 814.20(b)(8)(ii)-(iii)]					
	1.	Does the applicant list prior applications or state that there were no prior applications? (may be located in CDRH Coversheet Form FDA 3514, Section F) If the applicant lists prior applications, address the applicable questions below:			<input type="checkbox"/>		<input type="checkbox"/>
		a.	510(k) # _____		<input type="checkbox"/>	<input type="checkbox"/>	
		i.	If this device has been the subject of an NSE decision, does the HDE application take into account any concerns related to safety or probable benefit that were previously communicated during the review of the prior 510(k) or through 510(k) correspondence?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		b.	IDE # _____		<input type="checkbox"/>	<input type="checkbox"/>	
		i.	Have the data presented in the HDE taken into account any safety or probable benefit concerns (e.g., “future considerations”) previously communicated during the review of prior IDE(s) or through IDE correspondence?		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	PMA # _____		<input type="checkbox"/>	<input type="checkbox"/>	

Contains Nonbinding Recommendations

Inventory of Organizational and Administrative Elements
(Requirements per 21 CFR 814.112 unless otherwise indicated)

Check “Yes” if item is present, “N/A” if it is not needed and “Not Present” if it is not included but needed.

<ul style="list-style-type: none"> A “Not Present” answer may result in a “Refuse to File” decision. Each element on the checklist should be addressed within the application. An applicant may provide a rationale for omission for any criteria that are deemed not applicable. If a rationale is provided, the criteria is considered Present (“Yes”). An assessment of the rationale will be considered during the review of the application. 				Present		Not Present (No)
				Yes	N/A	
		i.	If a previously submitted PMA for this device has been withdrawn or denied, does the current HDE application take into account any issues related to safety or probable benefit raised during review of the prior PMA(s) or through PMA correspondence?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		d.	HDE # _____	<input type="checkbox"/>	<input type="checkbox"/>	
		i.	If a previously submitted HDE application for this device has been withdrawn or denied, does the current HDE application take into account any issues related to safety or probable benefit raised during review of the prior HDE application or through HDE correspondence?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		e.	Modular HDE # _____	<input type="checkbox"/>	<input type="checkbox"/>	
		i.	If “Yes”, how many modules submitted? _____ How many modules were closed? _____			
		ii.	If there are modules that are on hold, does the HDE address outstanding deficiencies?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
	2.	Does the applicant list Q-Submission(s) regarding the device or this application in which FDA feedback regarding data or information related to safety and/or probable benefit in the HDE was provided electronically or during a meeting (in person or by phone), or state that there were no prior Q-Submission interactions with the FDA regarding this application? If the applicant lists Q-Submissions, address the applicable questions below:		<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		a.	Q-Submission # _____ Meeting date(s), if applicable _____			
		b.	Copy of minutes from each meeting or other written feedback?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
		c.	Were all FDA concerns or action items previously presented to the applicant in the Q-Submission minutes or feedback addressed in the HDE or has the applicant provided a detailed scientific or clinical justification for an alternative approach?	<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Filing Decision Questions

A “No” answer will typically result in a Not-Filed decision.

		Yes	No
Decision 1	Is the HDE complete? If, on its face, the HDE is missing one or more required elements (identified above), such that the application is not sufficiently complete to permit substantive review, answer “No.”	<input type="checkbox"/>	<input type="checkbox"/>

Contains Nonbinding Recommendations

Filing Decision Questions			
A “No” answer will typically result in a Not-Filed decision.			
		Yes	No
Decision 2	<p>From only an administrative review, does the HDE include information that appears to constitute valid scientific evidence? ‡</p> <p>Only answer “No” if it is clear that the HDE is supported solely by information that 21 CFR 860.7 identifies as not constituting valid scientific evidence:</p> <ul style="list-style-type: none"> • isolated case reports • random experience • reports lacking sufficient details to permit scientific evaluation • unsubstantiated opinions <p>Comments:</p>	<input type="checkbox"/>	<input type="checkbox"/>
Decision 3	<p>Does the HDE address the key nonclinical and clinical issues identified by FDA prior to submission of the HDE application?</p> <p>OR</p> <p>Has the applicant provided a detailed scientific or clinical justification for the alternate approach?</p> <p>Section B of the checklist outlines questions intended to identify when the FDA has previously provided specific feedback to the applicant relevant to an evaluation of the risks and probable benefits of the device through one or more mechanisms, such as a prior HDE or PMA application, a prior “Not Substantially Equivalent” decision on a 510(k), Investigational Device Exemption (IDE) letters, feedback through the Q-submission Program, a Determination or Agreement meeting(s), or other substantive communication with FDA. If such information has been communicated to the applicant through one or more of these mechanisms, and the HDE application addresses each of the key nonclinical and clinical issues identified by FDA, the answer to the above question is “Yes.” Furthermore, if some of these key issues previously identified by FDA are not addressed, but the HDE application contains a scientific or clinical justification for the omission or an alternative approach, the answer to the above question is “Yes.” These cases do not preclude the responsible review Division from accepting the HDE application.</p> <p>In this context, the term “key issues” is meant to refer to issues that are central to FDA’s review of the device’s risks and probable benefit under sections 515 and 520(m) of the FD&C Act. Examples of key issues include: need for long-term nonclinical studies (e.g., biocompatibility, carcinogenicity, or other animal studies), and certain clinical study parameters (e.g., sample size, patient population, study design, and endpoints). These key issues are typically device-specific. As a result, the decision of FDA to “Refuse to File” an HDE application based on this criterion can only be made after carefully considering the following questions:</p> <p><i>Are the types of necessary nonclinical and clinical studies well-known in the scientific and medical communities for the particular device?</i></p> <p>For an “established” device type, the types of nonclinical and clinical studies that we would expect in a PMA are likely to be well-known both within FDA and in the scientific and medical communities and, as such, are often included as part of an FDA guidance document and/or consensus standard. You should bear in mind that, for HDEs, the device may not be of an established device type.</p>	<input type="checkbox"/>	<input type="checkbox"/>

[‡] For example, this information could be in the form of results of nonclinical laboratory studies with the device or results of clinical experience or investigations that are relevant to an assessment of the risks and probable benefits of the device.

Contains Nonbinding Recommendations

Filing Decision Questions			
A “No” answer will typically result in a Not-Filed decision.			
		Yes	No
	<p><i>Were the issues conveyed to the applicant as part of a documented regulatory process?</i></p> <p>Examples of a documented regulatory process include:</p> <ul style="list-style-type: none"> • interaction through the Q-submission process, • prior PMA or HDE application, • prior “Not Substantially Equivalent” decision on a 510(k), • IDE letters, or • letter(s) issued as a result of Determination or Agreement meetings. <p><i>Were the issues conveyed to the applicant related to insufficient effectiveness data?</i></p> <p>Devices approved under an HDE application are exempt from the requirement to demonstrate a reasonable assurance of effectiveness. If an issue that is not addressed in the current HDE application relates to insufficient effectiveness data, filing the HDE may be appropriate in cases for which accepting a PMA would not.</p> <p>FDA staff should only designate an HDE “Refuse to File” based on a “No” response to “Acceptance Decision 3” in instances where the key issues were identified by FDA staff as part of a documented regulatory process.</p>		

Digital Signature Concurrence Table	
Reviewer Sign-Off	
Team Lead/Assistant Director Sign-Off	
Division Sign-Off	
Office Sign-Off (for NOFI only)	

Appendix B – Considerations for the Probable Benefit-Risk Assessment

As discussed in Section VI of this guidance, FDA considers the same factors described in FDA’s benefit-risk framework for evaluating PMAs or De Novo requests when assessing probable benefits and risks for HDE applications. Refer to the FDA guidance document, “[Factors to Consider when Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications](https://www.fda.gov/regulatory-information/search-fda-guidance-documents/factors-consider-when-making-benefit-risk-determinations-medical-device-premarket-approval-and-de-novo-classifications)”,¹ for a description of those factors. It should be clearly noted, however, that probable benefit and probable benefit-risk determinations under an HDE are different from those under a PMA or a De Novo request. Please refer to Sections V and VI of the guidance for further discussion related to these differences and the probable benefit-risk assessment. The tools identified in Appendices B and C are meant to serve complementary roles, and both should be completed as part of the probable benefit-risk assessment.

Instructions for FDA Staff: You should make your recommendation regarding the probable benefit-risk assessment based on the totality of the evidence. The probable benefit-risk assessment is part of the decision whether to approve the application, but it does not include an assessment of all applicable requirements for approval. An indication from these tools that the probable benefits outweigh the probable risks does not mean that the application satisfies other applicable requirements for an HDE application.

The following questions are intended as a sequential method to help weigh various factors as part of the probable benefit-risk assessment. As such, the questions are intended to help identify and explain which factors and considerations are critical in making a probable benefit-risk assessment for a particular device. However, the questions are not intended to suggest that considerations other than those listed in the completed worksheet are irrelevant. This checklist should be used when non-clinical and/or clinical evidence has been submitted in the form of valid scientific evidence.

Consider questions 1-8 for the proposed Indications for Use, until you reach a recommendation either that the probable benefits outweigh the probable risks or to move to question 9, which prompts you to consider a narrowed Indications for Use. When considering an acceptable, narrowed Indications for Use, interact with the applicant to reach agreement on a narrowed Indications for Use. However, as reflected under question 1, if the evidence does not support a finding of probable benefit under the proposed Indications for Use (or narrowed Indications for Use), or evidence does not support a finding of probable benefit for the proposed Indications for Use and agreement on narrowed Indications for Use is not achievable or applicable, the application would not be approvable.

¹ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/factors-consider-when-making-benefit-risk-determinations-medical-device-premarket-approval-and-de>

Assessment of Probable Benefit

1. Is there any evidence of clinical benefit?

Note that in lieu of summaries, conclusions, and results of clinical investigations required under 21 CFR 814.20(b)(3)(v)(B), (b)(3)(vi), and (b)(6)(ii), HDE applicants are required to submit summaries, conclusions, and results of all clinical experience or investigations (whether adverse or supportive) reasonably obtainable by the applicant that are relevant to an assessment of the risks and probable benefits of the device (see 21 CFR 814.104(b)(4)(i)).

Is a probable clinical benefit demonstrated for the device for this indication (e.g., from any one or more of the primary and/or secondary datasets or from associated real-world evidence)? Probable benefit may be considered in terms of how a patient feels, functions, survives, or an acceptable surrogate outcome. This information may be collected using validated tools such as quality of life questionnaires, if appropriate. Probable benefit may also be considered in terms of convenience in managing or diagnosing a disease or condition. Probable benefit should be considered based on the assessment of the data. *Select any of the following that demonstrate probable benefit, and then answer the question in the box below.*

- ☐ A favorable change in at least 1 clinical assessment that is equal to or greater than seen in the control group
- ☐ A favorable change in at least 1 clinical assessment that meets a predetermined performance goal
- ☐ A favorable change in at least 1 clinical assessment that meets or surpasses a minimally important clinical difference
- ☐ A favorable change in at least 1 clinical assessment that is equal to or greater than changes seen with other available modalities for the disease or condition
- ☐ A favorable change that would be meaningful to patients considering the severity, chronicity, etc., of the condition, taking into consideration patient-reported outcomes (PRO) and health-related quality of life
- ☐ A favorable change in non-clinical data or modeling that is deemed to be predictive of clinical outcomes
- ☐ A favorable clinical performance characteristic (e.g., sensitivity/PPA², specificity/NPA³, etc.) for screening, diagnosis, prognosis, monitoring, or treatment selection
- ☐ Acceptable performance characteristics for analytical validation of the device
- ☐ Other(s)
- ☐ None

Question 1: Is there any evidence of probable clinical benefit?

- ☐ YES → Continue to Question 2
- ☐ NO → Move to Question 9

² PPA: Positive Percent Agreement

³ NPA: Negative Percent Agreement

Contains Nonbinding Recommendations

2. What is the extent of uncertainty for the probable benefits?

Recognizing that some extent of uncertainty always exists, select the sources of uncertainty, if applicable, in the data that affect your assessment of the clinical benefit. Consider sources of uncertainty related to clinical and/or analytical performance characteristics (e.g., sensitivity, specificity, accuracy, precision, reproducibility, as applicable). *Select any of the following that demonstrate sources of uncertainty for the probable benefits, and then answer the question in the box below.*

- ☐ Inconsistent or conflicting results between studies
- ☐ Wide confidence intervals surrounding the point estimate(s) and/or odds ratio(s)
- ☐ High subject or specimen loss-to-follow-up at critical assessment point(s)
- ☐ Large amount of missing data at critical assessment time(s) +/- imputation
- ☐ Significant number of major protocol deviations
- ☐ Impact of confounding interventions or physiological factors
- ☐ Inconsistent user experience or user experience not representative of likely real-world user
- ☐ Unclear correlation between non-clinical data, pre-selected enriched data, or computer modeling and clinical performance
- ☐ Surrogate endpoint has not been demonstrated to correlate with a clinical outcome
- ☐ Real-World Evidence (RWE) is not relevant or reliable for the purposes of the proposed analysis
- ☐ Inspectional findings
- ☐ Study design or results lead to lack of generalizability for the intended use population or specific clinical subpopulations
- ☐ Physiological or clinically meaningful range of the diagnostic output is unknown, or generalizability of proposed clinical cut-off is unknown
- ☐ Imperfect comparator method used to calculate performance characteristics
- ☐ Other(s)
- ☐ None

Question 2: What is the extent of uncertainty for the probable benefits?

- ☐ Low → Continue to Question 3; consider suggesting a different kind of marketing application.
- ☐ Med → Continue to Question 3
- ☐ High → Continue to Question 3

Summary of the Assessment of Probable Benefit

For the Proposed Indications for Use:

Considering responses to Questions 1 and 2, summarize the Assessment of Probable Benefit for the proposed Indications for Use. Include a description of your assessment of the extent of probable benefit, considering the type, magnitude, and probability of benefit(s); and the duration of effects. Include a description of the impact of uncertainty on your Assessment of Probable Benefit. If no benefit is identified, briefly explain why.

Assessment of Risk

3. Are known/probable risks more than minimal?

Select any of the following elements that demonstrate sources of known/probable risks that are more than minimal, and then answer the question in the box below.

- ☐ Adverse events (AEs) or outcomes related to the device itself
- ☐ AEs or outcomes related to the use of the device or procedure to use the device
- ☐ AEs or outcomes related to anesthesia or sedation to use the device
- ☐ AEs or outcomes due to subsequent tests/treatments needed (e.g., radiation from CT scans)
- ☐ AEs or outcomes, not seen in the study/data, but probable based on “class effect” or events known to occur with similar technologies
- ☐ False positive/false negative/failed to provide a result for diagnostics
- ☐ Treatment or diagnostic intended to be used as a standalone rather than an adjunctive use
- ☐ Other(s)
- ☐ None

Question 3: Are known/probable risks more than minimal?

- ☐ YES → Continue to Question 4
- ☐ NO → Continue to Question 4

4. What is the extent of uncertainty for the probable risks?

Recognizing that some extent of uncertainty always exists, select the sources of uncertainty, if applicable, in the data regarding the adverse events/outcomes or risks. Select any of the following that demonstrate sources of uncertainty for the probable risks, and then answer the question in the box below.

- ☐ Insufficient patient numbers to detect serious events or false positives/false negatives
- ☐ Insufficient duration of follow-up to detect delayed/late events
- ☐ Lack of data on repeated exposure to the device/use
- ☐ Inconsistent or conflicting results between studies
- ☐ Proper evaluations not performed as part of the study protocol to adequately detect certain AEs
- ☐ Poor or inconsistent adverse event definitions and documentation
- ☐ Events likely confounded by, and attributed to, other comorbidities or treatment modalities
- ☐ High subject loss-to-follow-up at critical assessment point(s)
- ☐ Large amount of missing data at critical assessment time(s) +/- imputation
- ☐ Significant number of major protocol deviations
- ☐ Inconsistent user experience or user experience not representative of likely real-world user
- ☐ Concerns related to performance characteristics (e.g., sensitivity/PPA, specificity/NPA)
- ☐ Imperfect comparator method used to calculate performance characteristics
- ☐ Other(s)
- ☐ None

Contains Nonbinding Recommendations

Question 4: What is the extent of uncertainty for the probable risks?

- ☐ Low → Continue to Question 5
- ☐ Med → Continue to Question 5
- ☐ High → Continue to Question 5

Summary of the Assessment of Probable Risk

If you answered “No” to Question 3 but “High” to Question 4, please explain here.

For the Proposed Indications for Use:

Summarize the Assessment of Probable Risk for the proposed Indications for Use. Include a description of your assessment of the extent of probable risk considering the severity, types, number and rates of harmful events associated with use of the device; probability of a harmful event; duration of harmful events; and risk from false-positive or false-negative results for diagnostics. Include a description of the impact of uncertainty on your Assessment of Probable Risk.

Assessment of Probable Benefit-Risk

Instructions for FDA Staff: Provide a recommendation based on the totality of the evidence. As noted above, the probable benefit-risk assessment is part of the decision regarding whether to approve an HDE application but is not an assessment of all applicable requirements.

To approve an HDE application, FDA must make, among other things, a determination that the device will not expose patients to an unreasonable or significant risk of illness or injury and that the probable benefit to health from the use of the device outweighs the risk of injury or illness from its use taking into account the probable benefits and risks of currently available devices or alternative forms of treatment.

If you answer “yes” for any Q5-8, explain your rationale for how the probable benefits outweigh the probable risks. You should also consider and recommend actions that would enhance the probable benefit-risk profile of the device, such as modifications to the proposed labeling, which may include additional appropriate warnings, and precautions, instructions for use, or presentation of data to help ensure the product labeling is transparent with respect to the probable benefits and risks.

If you answer “unable to conclude” for Q5-8, please provide the information that you believe would be needed to support a determination that probable benefits outweigh the probable risks for the Indications for Use under consideration in the summary text boxes and also proceed to Q9.

Question 5: Do the Probable Benefits outweigh the Risks, considering the assessment of the Probable Benefit and Risk and the extent of uncertainty identified above?

- ☐ Yes – The probable benefits outweigh the risks such that, for this device, additional consideration of relevant factors would not change that determination.
- ☐ Unable to conclude that probable benefits outweigh the risks – further discussion and consideration of relevant factors is appropriate → Continue to Question 6

Contains Nonbinding Recommendations

Summary of the Assessment of Probable Benefit-Risk

For the Proposed Indications for Use:

Summarize the probable benefit(s) that have been demonstrated for the proposed Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how available alternative modalities, including their probable benefits and risks, affect your assessment. Include a description of how uncertainty regarding Probable Benefit(s) and Risk(s) affects your assessment.

6. Do the Probable Benefits outweigh the Risks, when taking into account the following additional considerations? *Select relevant considerations, and then answer the question in the box below.*

- ☐ Understanding of patient willingness or unwillingness to accept a large extent of uncertainty in the probable benefits and/or risks
- ☐ Available patient preference information (PPI) showing patient willingness or unwillingness to accept the probable risks in exchange for the probable benefits. In circumstances where it is not feasible to obtain PPI (e.g., some pediatric or impaired patient populations), care-partner perspectives may be considered.
- ☐ Understanding of care-partner perspectives on the probable benefits and risks for a device where applicable (e.g., ease of care that may affect patient care and outcomes)
- ☐ Available qualitative or quantitative PPI on the relative desirability or acceptability to patients of outcomes or other attributes that differ among alternative health interventions
- ☐ Understanding that the device represents novel technology for which the current device technology is different
- ☐ Ability to manage or diagnose the condition and consideration of natural history of disease progression in the absence of the intervention or diagnostic information with the device under review
- ☐ The device avoids serious harm associated with available therapies for the disease or condition
- ☐ The adverse events associated with use of the device are reversible
- ☐ Type of intervention required to address the harmful event (e.g., medication, surgery)
- ☐ Understanding of mechanistic plausibility and/or “class effect” (e.g., familiarity with similar technology)
- ☐ Other(s)
- ☐ None

Question 6: Do the Probable Benefits outweigh the Risks, when taking into account additional relevant considerations?

- ☐ Yes – The probable benefits outweigh the risks such that, for this device, additional consideration of relevant factors would not change that determination.
- ☐ Unable to conclude that probable benefits outweigh the risks – further discussion and consideration of risk mitigation measures is appropriate → Continue to Question 7

Contains Nonbinding Recommendations

Summary of the Assessment of Probable Benefit-Risk, taking into account additional relevant considerations

For the Proposed Indications for Use:

Summarize the probable benefit(s) that have been demonstrated for the proposed Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how available alternative modalities, including their probable benefits and risks, affect your assessment. Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment. Include a description of how patient perspectives affected your assessment.

- 7. Can the risks be mitigated, so that Probable Benefits outweigh the Risks? Consider if the Probable Benefits outweigh the Risks if risk mitigation strategies are incorporated to lower the probability of a harmful event occurring and improve the probable benefit-risk profile of the device. Select relevant considerations, and then answer the question in the box below.**

- ☐ Additional descriptions of known and probable benefits and risks in physician and patient labeling, including appropriate Contraindications, Warnings, and Precautions and description of the clinical events
- ☐ Additional warnings noting limitations of safety information (e.g., “The safety of the use of this device in [situation] has not been evaluated.”)
- ☐ Labeling the device for prescription use only

Training:

- ☐ Limitation to caregivers with certain qualifications or clinical training
- ☐ Limit to users with a minimum set of qualifications and/or training
- ☐ Physician/user training program

Other:

- ☐ Device tracking
- ☐ Other(s)
- ☐ None

Question 7: Can the risks be mitigated, so that Probable Benefits outweigh the Risks?

- ☐ Yes – The probable benefits outweigh the risks such that, for this device, additional consideration of relevant factors would not change that determination.
- ☐ Unable to conclude that probable benefits outweigh the risks – further discussion and consideration of postmarket actions is appropriate → Continue to Question 8

Summary of the Assessment of Probable Benefit-Risk, considering risk mitigation strategies

For the Proposed Indications for Use:

Summarize the probable benefit(s) that have been demonstrated for the proposed Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how available alternative modalities, including their probable benefits and risks, affect your assessment. Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment. Include a description of how patient perspectives affected your assessment.

8. Do the Probable Benefits outweigh the Risks considering the use of postmarket actions? *Select appropriate postmarket action(s), and then answer the question in the box below.*

- ☐ Collection of additional and/or confirmatory non-clinical performance data in the postmarket space (e.g., post-approval study, postmarket surveillance)
- ☐ Collection of additional and/or confirmatory clinical data in the postmarket space (e.g., post-approval study, postmarket surveillance)

If either non-clinical or clinical performance data collections in the postmarket space are checked, consider:

- ☐ The feasibility of postmarket data collection and likelihood that postmarket data collection will be completed within a reasonable timeframe
- ☐ Whether it would be appropriate for labeling to include description of postmarket data collection and its purpose

- ☐ Other(s)
- ☐ None

Question 8: Do the Probable Benefits outweigh the Risks considering the use of postmarket actions?

- ☐ Yes – The probable benefits outweigh the risks.
- ☐ Unable to conclude that probable benefits outweigh the risks → Continue to Question 9

Contains Nonbinding Recommendations

Summary of the Assessment of Probable Benefit-Risk, considering postmarket actions

For the Proposed Indications for Use:

Summarize the probable benefits(s) that have been demonstrated for the proposed Indications for Use and your assessment of how Probable Benefit(s) compare to Risks. Include a description of how available alternative modalities, including their probable benefits and risks, affect your assessment. Include a description of how uncertainty regarding Probable Benefit(s) and Risks affects your assessment. Include a description of how patient perspectives affected your assessment.

Question 9: Is there any evidence of probable clinical benefit for a narrowed Indications for Use?

- ☐ Yes → Return to Question 1 and proceed with narrowed Indications for Use
- ☐ No → Do not approve the application

Appendix C – Probable Benefit-Risk Assessment Summary

HDE Probable Benefit-Risk Assessment Summary HDE Questions Based on the totality of the data Device Name: HDE Number:	Proposed Indications for Use.
Assessment of Probable Benefit	Considering benefit in terms of <ul style="list-style-type: none"> Type Magnitude Probability Duration of effects Patient perspective (or care-partner, if applicable)
1. Is there any evidence of clinical benefit?	<input type="checkbox"/> YES → Q2 <input type="checkbox"/> NO → Do not approve for proposed Indications for Use; proceed to Q9
2. What is the extent of uncertainty for the Probable Benefits? ¹	<input type="checkbox"/> High <input type="checkbox"/> Medium <input type="checkbox"/> Low Continue to Q3
Assessment of Risk	Considering risk in terms of <ul style="list-style-type: none"> Severity, types, number and rates of harmful events Probability of a harmful event Duration of harmful events Risks from false-positive or false-negative results Patient perspective (or care-partner, if applicable)
3. Are known/probable risks more than minimal?	<input type="checkbox"/> YES → Q4 <input type="checkbox"/> NO → Q4
4. What is the extent of uncertainty for the Probable Risks?	<input type="checkbox"/> High <input type="checkbox"/> Med <input type="checkbox"/> Low Continue to Q5
Assessment of Probable Benefit-Risk	
5. Do the Probable Benefits outweigh the Risks? ²	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Unable to conclude that probable benefits outweigh the risks → Q6
6. Do the Probable Benefits outweigh the Risks, taking into account additional considerations?	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Unable to conclude that probable benefits outweigh the risks → Q7
7. Can the risks be mitigated, so that Probable Benefits outweigh the Risks?	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Unable to conclude that probable benefits outweigh the risks → Q8
8. Do the Probable Benefits outweigh the Risks considering the use of postmarket actions?	<input type="checkbox"/> YES → Worksheet complete <input type="checkbox"/> Unable to conclude that probable benefits outweigh the risks → Q9

¹ Instructions: If the extent of uncertainty is low, then consider whether a different kind of marketing application would be appropriate. However, low uncertainty does not necessarily imply clinically meaningful benefit.

² Instructions: For an HDE, take into account the probable benefits and risks of currently available devices or alternative forms of treatment.

9. Is there any evidence of clinical benefit for a narrowed Indications for Use?	<input type="checkbox"/> YES → Return to Q1 and proceed with narrowed Indications for Use ³ <input type="checkbox"/> NO → Do not approve the application
--	--

³ Instructions: The term “indications for use” describes the disease or condition that the device will diagnose, treat, prevent, cure, or mitigate, including a description of the patient population for which the device is intended. See 21 CFR 814.20(b)(3)(i) and 814.104(b)(4). Consider the probable benefits and risks for a modified population for the proposed use, a modified indication for the proposed population, or both a modified indication and modified population, which would translate into a ‘narrowing’ of the Indications for Use from what was originally proposed. Note that probable benefit and probable benefit-risk determinations for HDEs are different from those under PMAs. For more information, refer to Section VI of this guidance.